

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc **SDG No.:** Q1780
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1780 **SAS No.:** Q1780
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	-4.96	2.0	248	-18	22	04/10/2025	10:37	LB135373
	Iron	107000	101000	106	85600	116500	04/10/2025	10:37	LB135373
	Lead	7.32			-12	12	04/10/2025	10:37	LB135373
ICSAB01	Copper	500	511	98	434	588	04/10/2025	10:42	LB135373
	Iron	93600	99300	94	84400	114500	04/10/2025	10:42	LB135373
	Lead	57.3	49.0	117	37	61	04/10/2025	10:42	LB135373
ICSA	Copper	-3.40	2.0	170	-18	22	04/10/2025	10:46	LB135373
	Iron	105000	101000	104	85600	116500	04/10/2025	10:46	LB135373
	Lead	-3.88			-12	12	04/10/2025	10:46	LB135373
ICSAB	Copper	533	511	104	434	588	04/10/2025	10:50	LB135373
	Iron	103000	99300	104	84400	114500	04/10/2025	10:50	LB135373
	Lead	33.0	49.0	67	37	61	04/10/2025	10:50	LB135373