

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Environmental, Inc **SDG No.:** Q1780
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1780 **SAS No.:** Q1780

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167554BS							
Copper	ug/L	150	148		99	85 - 115	P
Iron	ug/L	1500	1370		91	85 - 115	P
Lead	ug/L	500	448		90	85 - 115	P