

## **CASE NARRATIVE**

**Lockwood, Kessler & Bartlett, Inc.**

**Project Name: Ansonia Landfill 2025**

**Project # N/A**

**Chemtech Project # Q1782**

**Test Name: VOCMS Group1**

### **A. Number of Samples and Date of Receipt:**

9 Water samples were received on 04/10/2025.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: Alkalinity, Ammonia, Anions Group1, BOD5, Dissolved Metals Group5, Metals Group4, pH, Redox Potential, TDS, TKN, Total Nitrogen, TSS, Turbidity and VOCMS Group1. This data package contains results for VOCMS Group1.

### **C. Analytical Techniques:**

The analysis performed on instrument MSVOA\_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI The analysis of VOCMS Group1 was based on method 8260-Low.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate for { VX0411WBSD01 } with File ID: VX045734.D met requirements for all samples except for Bromochloromethane[129%] but associated samples have not positive hit for this compound therefore no corrective action was taken. The Blank analysis did not indicate the presence of lab contamination.

The %RSD is greater than 20% in the Initial Calibration method (82X040225W.M) for t-1,3dichloropropene, this compound is passing on Linear Regression.

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

### **E. Additional Comments:**

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Trip Blank was not provided with this set of samples.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_