

Preparation Blank Summary

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q1782

Project: Ansonia Landfill 2025

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB135377BL							
TDS	mg/L	< 5.0000	5.0000	U	1.0	10	04/11/2025
Sample ID: LB135389BLW							
Bromide	mg/L	< 1.0000	1.0000	U	0.37	2	04/11/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.19	0.6	04/11/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.11	0.4	04/11/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.074	0.6	04/11/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.095	0.5	04/11/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.46	3	04/11/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.34	1	04/11/2025
Sample ID: LB135390BL							
BOD5	mg/L	< 0.2000	0.2000	U	0.20	2.0	04/11/2025
Sample ID: LB135393BL							
Turbidity	NTU	0.446	0.5000	J	0.15	1.0	04/11/2025
Sample ID: LB135430BL							
TSS	mg/L	1	2.0000	J	1	4	04/15/2025
Sample ID: LB135460BLW							
Alkalinity	mg/L	< 1.0000	1.0000	U	1	2	04/16/2025
Sample ID: PB167617BL							
TKN	mg/L	< 0.2500	0.2500	U	0.11	0.5	04/18/2025
Sample ID: PB167618BL							
Ammonia as N	mg/L	< 0.0500	0.0500	U	0.03	0.1	04/18/2025