

LAB CHRONICLE

OrderID:	Q1782	OrderDate:	4/10/2025 1:22:00 PM
Client:	Lockwood, Kessler & Bartlett, Inc.	Project:	Ansonia Landfill 2025
Contact:	John Gerlach	Location:	K11,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1782-01	MW-1	Water	Metals Group4	6010D	04/09/25	04/14/25	04/15/25	04/10/25
Q1782-02	MW-1	Water	Dissolved Metals Group5	6010D	04/09/25	04/14/25	04/15/25	04/10/25
Q1782-03	MW-2	Water	Metals Group4	6010D	04/09/25	04/14/25	04/15/25	04/10/25
Q1782-04	MW-2	Water	Dissolved Metals Group5	6010D	04/09/25	04/14/25	04/15/25	04/10/25
Q1782-05	MW-3	Water	Metals Group4	6010D	04/09/25	04/14/25	04/15/25	04/10/25
Q1782-06	MW-3	Water	Dissolved Metals Group5	6010D	04/09/25	04/14/25	04/15/25	04/10/25
Q1782-07	MW-4	Water	Metals Group4	6010D	04/09/25	04/14/25	04/15/25	04/10/25
Q1782-08	MW-4	Water	Dissolved Metals Group5	6010D	04/09/25	04/14/25	04/15/25	04/10/25

Hit Summary Sheet SW-846

SDG No.: Q1782 **Order ID:** Q1782
Client: Lockwood, Kessler & Bartlett, Inc. **Project ID:** Ansonia Landfill 2025

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : MW-1								
Q1782-01	MW-1	Water	Barium	43.9	J	6.28	50.0	ug/L
Q1782-01	MW-1	Water	Cadmium	1.18	J	0.094	3.00	ug/L
Q1782-01	MW-1	Water	Chromium	1.67	J	0.66	5.00	ug/L
Q1782-01	MW-1	Water	Nickel	4.15	J	0.85	20.0	ug/L
Q1782-01	MW-1	Water	Potassium	4240		685	1000	ug/L
Q1782-01	MW-1	Water	Sodium	63400		237	1000	ug/L
Q1782-01	MW-1	Water	Zinc	129		1.75	20.0	ug/L
Client ID : MW-1								
Q1782-02	MW-1	Water	Iron	158		18.5	50.0	ug/L
Q1782-02	MW-1	Water	Manganese	102		1.46	10.0	ug/L
Client ID : MW-2								
Q1782-03	MW-2	Water	Barium	115		6.28	50.0	ug/L
Q1782-03	MW-2	Water	Cadmium	0.18	J	0.094	3.00	ug/L
Q1782-03	MW-2	Water	Chromium	10.9		0.66	5.00	ug/L
Q1782-03	MW-2	Water	Copper	16.0		7.07	10.0	ug/L
Q1782-03	MW-2	Water	Nickel	6.76	J	0.85	20.0	ug/L
Q1782-03	MW-2	Water	Potassium	16500		685	1000	ug/L
Q1782-03	MW-2	Water	Sodium	192000		237	1000	ug/L
Q1782-03	MW-2	Water	Zinc	46.9		1.75	20.0	ug/L
Client ID : MW-2								
Q1782-04	MW-2	Water	Iron	113		18.5	50.0	ug/L
Q1782-04	MW-2	Water	Manganese	55.6		1.46	10.0	ug/L
Client ID : MW-3								
Q1782-05	MW-3	Water	Barium	117		6.28	50.0	ug/L
Q1782-05	MW-3	Water	Potassium	12400		685	1000	ug/L
Q1782-05	MW-3	Water	Sodium	205000		237	1000	ug/L
Q1782-05	MW-3	Water	Zinc	9.44	J	1.75	20.0	ug/L
Client ID : MW-3								
Q1782-06	MW-3	Water	Iron	74.5		18.5	50.0	ug/L
Q1782-06	MW-3	Water	Manganese	51.2		1.46	10.0	ug/L
Client ID : MW-4								
Q1782-07	MW-4	Water	Barium	113		6.28	50.0	ug/L
Q1782-07	MW-4	Water	Potassium	12200		685	1000	ug/L

Hit Summary Sheet
SW-846

SDG No.: Q1782

Order ID: Q1782

Client: Lockwood, Kessler & Bartlett, Inc.

Project ID: Ansonia Landfill 2025

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q1782-07	MW-4	Water	Sodium	202000		237	1000	ug/L
Q1782-07	MW-4	Water	Zinc	9.56	J	1.75	20.0	ug/L
Client ID : MW-4								
Q1782-08	MW-4	Water	Iron	71.3		18.5	50.0	ug/L
Q1782-08	MW-4	Water	Manganese	32.7		1.46	10.0	ug/L



SAMPLE DATA

Report of Analysis

Client:	Lockwood, Kessler & Bartlett, Inc.	Date Collected:	04/09/25
Project:	Ansonia Landfill 2025	Date Received:	04/10/25
Client Sample ID:	MW-1	SDG No.:	Q1782
Lab Sample ID:	Q1782-01	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-36-0	Antimony	2.06	U	1	2.06	25.0	ug/L	04/14/25 10:30	04/15/25 19:50	SW6010	SW3010
7440-38-2	Arsenic	3.48	U	1	3.48	10.0	ug/L	04/14/25 10:30	04/15/25 19:50	SW6010	SW3010
7440-39-3	Barium	43.9	J	1	6.28	50.0	ug/L	04/14/25 10:30	04/15/25 19:50	SW6010	SW3010
7440-41-7	Beryllium	0.13	U	1	0.13	3.00	ug/L	04/14/25 10:30	04/15/25 19:50	SW6010	SW3010
7440-43-9	Cadmium	1.18	J	1	0.094	3.00	ug/L	04/14/25 10:30	04/15/25 19:50	SW6010	SW3010
7440-47-3	Chromium	1.67	J	1	0.66	5.00	ug/L	04/14/25 10:30	04/15/25 19:50	SW6010	SW3010
7440-48-4	Cobalt	0.50	U	1	0.50	15.0	ug/L	04/14/25 10:30	04/15/25 19:50	SW6010	SW3010
7440-50-8	Copper	7.07	U	1	7.07	10.0	ug/L	04/14/25 10:30	04/15/25 19:50	SW6010	SW3010
7439-92-1	Lead	3.51	U	1	3.51	6.00	ug/L	04/14/25 10:30	04/15/25 19:50	SW6010	SW3010
7440-02-0	Nickel	4.15	J	1	0.85	20.0	ug/L	04/14/25 10:30	04/15/25 19:50	SW6010	SW3010
7440-09-7	Potassium	4240		1	685	1000	ug/L	04/14/25 10:30	04/15/25 19:50	SW6010	SW3010
7782-49-2	Selenium	5.88	U	1	5.88	10.0	ug/L	04/14/25 10:30	04/15/25 19:50	SW6010	SW3010
7440-22-4	Silver	0.58	U	1	0.58	5.00	ug/L	04/14/25 10:30	04/15/25 19:50	SW6010	SW3010
7440-23-5	Sodium	63400		1	237	1000	ug/L	04/14/25 10:30	04/15/25 19:50	SW6010	SW3010
7440-28-0	Thallium	2.32	U	1	2.32	20.0	ug/L	04/14/25 10:30	04/15/25 19:50	SW6010	SW3010
7440-62-2	Vanadium	3.06	U	1	3.06	20.0	ug/L	04/14/25 10:30	04/15/25 19:50	SW6010	SW3010
7440-66-6	Zinc	129		1	1.75	20.0	ug/L	04/14/25 10:30	04/15/25 19:50	SW6010	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group4			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Lockwood, Kessler & Bartlett, Inc.	Date Collected:	04/09/25
Project:	Ansonia Landfill 2025	Date Received:	04/10/25
Client Sample ID:	MW-1	SDG No.:	Q1782
Lab Sample ID:	Q1782-02	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7439-89-6	Iron	158		1	18.5	50.0	ug/L	04/14/25 10:30	04/15/25 19:54	SW6010	SW3010
7439-96-5	Manganese	102	N*	1	1.46	10.0	ug/L	04/14/25 10:30	04/15/25 19:54	SW6010	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Dissolved Metals Group5			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Lockwood, Kessler & Bartlett, Inc.	Date Collected:	04/09/25
Project:	Ansonia Landfill 2025	Date Received:	04/10/25
Client Sample ID:	MW-2	SDG No.:	Q1782
Lab Sample ID:	Q1782-03	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-36-0	Antimony	2.06	U	1	2.06	25.0	ug/L	04/14/25 10:30	04/15/25 20:07	SW6010	SW3010
7440-38-2	Arsenic	3.48	U	1	3.48	10.0	ug/L	04/14/25 10:30	04/15/25 20:07	SW6010	SW3010
7440-39-3	Barium	115		1	6.28	50.0	ug/L	04/14/25 10:30	04/15/25 20:07	SW6010	SW3010
7440-41-7	Beryllium	0.13	U	1	0.13	3.00	ug/L	04/14/25 10:30	04/15/25 20:07	SW6010	SW3010
7440-43-9	Cadmium	0.18	J	1	0.094	3.00	ug/L	04/14/25 10:30	04/15/25 20:07	SW6010	SW3010
7440-47-3	Chromium	10.9		1	0.66	5.00	ug/L	04/14/25 10:30	04/15/25 20:07	SW6010	SW3010
7440-48-4	Cobalt	0.50	U	1	0.50	15.0	ug/L	04/14/25 10:30	04/15/25 20:07	SW6010	SW3010
7440-50-8	Copper	16.0		1	7.07	10.0	ug/L	04/14/25 10:30	04/15/25 20:07	SW6010	SW3010
7439-92-1	Lead	3.51	U	1	3.51	6.00	ug/L	04/14/25 10:30	04/15/25 20:07	SW6010	SW3010
7440-02-0	Nickel	6.76	J	1	0.85	20.0	ug/L	04/14/25 10:30	04/15/25 20:07	SW6010	SW3010
7440-09-7	Potassium	16500		1	685	1000	ug/L	04/14/25 10:30	04/15/25 20:07	SW6010	SW3010
7782-49-2	Selenium	5.88	U	1	5.88	10.0	ug/L	04/14/25 10:30	04/15/25 20:07	SW6010	SW3010
7440-22-4	Silver	0.58	U	1	0.58	5.00	ug/L	04/14/25 10:30	04/15/25 20:07	SW6010	SW3010
7440-23-5	Sodium	192000		1	237	1000	ug/L	04/14/25 10:30	04/15/25 20:07	SW6010	SW3010
7440-28-0	Thallium	2.32	U	1	2.32	20.0	ug/L	04/14/25 10:30	04/15/25 20:07	SW6010	SW3010
7440-62-2	Vanadium	3.06	U	1	3.06	20.0	ug/L	04/14/25 10:30	04/15/25 20:07	SW6010	SW3010
7440-66-6	Zinc	46.9		1	1.75	20.0	ug/L	04/14/25 10:30	04/15/25 20:07	SW6010	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group4			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Lockwood, Kessler & Bartlett, Inc.	Date Collected:	04/09/25
Project:	Ansonia Landfill 2025	Date Received:	04/10/25
Client Sample ID:	MW-2	SDG No.:	Q1782
Lab Sample ID:	Q1782-04	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7439-89-6	Iron	113		1	18.5	50.0	ug/L	04/14/25 10:30	04/15/25 20:11	SW6010	SW3010
7439-96-5	Manganese	55.6	N*	1	1.46	10.0	ug/L	04/14/25 10:30	04/15/25 20:11	SW6010	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Dissolved Metals Group5			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Lockwood, Kessler & Bartlett, Inc.	Date Collected:	04/09/25
Project:	Ansonia Landfill 2025	Date Received:	04/10/25
Client Sample ID:	MW-3	SDG No.:	Q1782
Lab Sample ID:	Q1782-05	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-36-0	Antimony	2.06	U	1	2.06	25.0	ug/L	04/14/25 10:30	04/15/25 20:16	SW6010	SW3010
7440-38-2	Arsenic	3.48	U	1	3.48	10.0	ug/L	04/14/25 10:30	04/15/25 20:16	SW6010	SW3010
7440-39-3	Barium	117		1	6.28	50.0	ug/L	04/14/25 10:30	04/15/25 20:16	SW6010	SW3010
7440-41-7	Beryllium	0.13	U	1	0.13	3.00	ug/L	04/14/25 10:30	04/15/25 20:16	SW6010	SW3010
7440-43-9	Cadmium	0.094	U	1	0.094	3.00	ug/L	04/14/25 10:30	04/15/25 20:16	SW6010	SW3010
7440-47-3	Chromium	0.66	U	1	0.66	5.00	ug/L	04/14/25 10:30	04/15/25 20:16	SW6010	SW3010
7440-48-4	Cobalt	0.50	U	1	0.50	15.0	ug/L	04/14/25 10:30	04/15/25 20:16	SW6010	SW3010
7440-50-8	Copper	7.07	U	1	7.07	10.0	ug/L	04/14/25 10:30	04/15/25 20:16	SW6010	SW3010
7439-92-1	Lead	3.51	U	1	3.51	6.00	ug/L	04/14/25 10:30	04/15/25 20:16	SW6010	SW3010
7440-02-0	Nickel	0.85	U	1	0.85	20.0	ug/L	04/14/25 10:30	04/15/25 20:16	SW6010	SW3010
7440-09-7	Potassium	12400		1	685	1000	ug/L	04/14/25 10:30	04/15/25 20:16	SW6010	SW3010
7782-49-2	Selenium	5.88	U	1	5.88	10.0	ug/L	04/14/25 10:30	04/15/25 20:16	SW6010	SW3010
7440-22-4	Silver	0.58	U	1	0.58	5.00	ug/L	04/14/25 10:30	04/15/25 20:16	SW6010	SW3010
7440-23-5	Sodium	205000		1	237	1000	ug/L	04/14/25 10:30	04/15/25 20:16	SW6010	SW3010
7440-28-0	Thallium	2.32	U	1	2.32	20.0	ug/L	04/14/25 10:30	04/15/25 20:16	SW6010	SW3010
7440-62-2	Vanadium	3.06	U	1	3.06	20.0	ug/L	04/14/25 10:30	04/15/25 20:16	SW6010	SW3010
7440-66-6	Zinc	9.44	J	1	1.75	20.0	ug/L	04/14/25 10:30	04/15/25 20:16	SW6010	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group4			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N = Spiked sample recovery not within control limits

Report of Analysis

Client:	Lockwood, Kessler & Bartlett, Inc.	Date Collected:	04/09/25
Project:	Ansonia Landfill 2025	Date Received:	04/10/25
Client Sample ID:	MW-3	SDG No.:	Q1782
Lab Sample ID:	Q1782-06	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7439-89-6	Iron	74.5		1	18.5	50.0	ug/L	04/14/25 10:30	04/15/25 20:21	SW6010	SW3010
7439-96-5	Manganese	51.2	N*	1	1.46	10.0	ug/L	04/14/25 10:30	04/15/25 20:21	SW6010	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Dissolved Metals Group5			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Lockwood, Kessler & Bartlett, Inc.	Date Collected:	04/09/25
Project:	Ansonia Landfill 2025	Date Received:	04/10/25
Client Sample ID:	MW-4	SDG No.:	Q1782
Lab Sample ID:	Q1782-07	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-36-0	Antimony	2.06	U	1	2.06	25.0	ug/L	04/14/25 10:30	04/15/25 20:25	SW6010	SW3010
7440-38-2	Arsenic	3.48	U	1	3.48	10.0	ug/L	04/14/25 10:30	04/15/25 20:25	SW6010	SW3010
7440-39-3	Barium	113		1	6.28	50.0	ug/L	04/14/25 10:30	04/15/25 20:25	SW6010	SW3010
7440-41-7	Beryllium	0.13	U	1	0.13	3.00	ug/L	04/14/25 10:30	04/15/25 20:25	SW6010	SW3010
7440-43-9	Cadmium	0.094	U	1	0.094	3.00	ug/L	04/14/25 10:30	04/15/25 20:25	SW6010	SW3010
7440-47-3	Chromium	0.66	U	1	0.66	5.00	ug/L	04/14/25 10:30	04/15/25 20:25	SW6010	SW3010
7440-48-4	Cobalt	0.50	U	1	0.50	15.0	ug/L	04/14/25 10:30	04/15/25 20:25	SW6010	SW3010
7440-50-8	Copper	7.07	U	1	7.07	10.0	ug/L	04/14/25 10:30	04/15/25 20:25	SW6010	SW3010
7439-92-1	Lead	3.51	U	1	3.51	6.00	ug/L	04/14/25 10:30	04/15/25 20:25	SW6010	SW3010
7440-02-0	Nickel	0.85	U	1	0.85	20.0	ug/L	04/14/25 10:30	04/15/25 20:25	SW6010	SW3010
7440-09-7	Potassium	12200		1	685	1000	ug/L	04/14/25 10:30	04/15/25 20:25	SW6010	SW3010
7782-49-2	Selenium	5.88	U	1	5.88	10.0	ug/L	04/14/25 10:30	04/15/25 20:25	SW6010	SW3010
7440-22-4	Silver	0.58	U	1	0.58	5.00	ug/L	04/14/25 10:30	04/15/25 20:25	SW6010	SW3010
7440-23-5	Sodium	202000		1	237	1000	ug/L	04/14/25 10:30	04/15/25 20:25	SW6010	SW3010
7440-28-0	Thallium	2.32	U	1	2.32	20.0	ug/L	04/14/25 10:30	04/15/25 20:25	SW6010	SW3010
7440-62-2	Vanadium	3.06	U	1	3.06	20.0	ug/L	04/14/25 10:30	04/15/25 20:25	SW6010	SW3010
7440-66-6	Zinc	9.56	J	1	1.75	20.0	ug/L	04/14/25 10:30	04/15/25 20:25	SW6010	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group4			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	Lockwood, Kessler & Bartlett, Inc.	Date Collected:	04/09/25
Project:	Ansonia Landfill 2025	Date Received:	04/10/25
Client Sample ID:	MW-4	SDG No.:	Q1782
Lab Sample ID:	Q1782-08	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7439-89-6	Iron	71.3		1	18.5	50.0	ug/L	04/14/25 10:30	04/15/25 20:59	SW6010	SW3010
7439-96-5	Manganese	32.7	N*	1	1.46	10.0	ug/L	04/14/25 10:30	04/15/25 20:59	SW6010	SW3010

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Dissolved Metals Group5			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q1782

Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q1782 **SAS No.:** Q1782

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Antimony	1050	1000	105	90 - 110	P	04/15/2025	15:47	lb135445
	Arsenic	1020	1000	102	90 - 110	P	04/15/2025	15:47	lb135445
	Barium	548	520	105	90 - 110	P	04/15/2025	15:47	lb135445
	Beryllium	488	510	96	90 - 110	P	04/15/2025	15:47	lb135445
	Cadmium	493	510	97	90 - 110	P	04/15/2025	15:47	lb135445
	Chromium	506	520	97	90 - 110	P	04/15/2025	15:47	lb135445
	Cobalt	478	520	92	90 - 110	P	04/15/2025	15:47	lb135445
	Copper	508	510	100	90 - 110	P	04/15/2025	15:47	lb135445
	Iron	9690	10000	97	90 - 110	P	04/15/2025	15:47	lb135445
	Lead	973	1000	97	90 - 110	P	04/15/2025	15:47	lb135445
	Manganese	489	520	94	90 - 110	P	04/15/2025	15:47	lb135445
	Nickel	480	530	90	90 - 110	P	04/15/2025	15:47	lb135445
	Potassium	10100	9900	102	90 - 110	P	04/15/2025	15:47	lb135445
	Selenium	1070	1000	107	90 - 110	P	04/15/2025	15:47	lb135445
	Silver	247	250	99	90 - 110	P	04/15/2025	15:47	lb135445
	Sodium	10800	10000	108	90 - 110	P	04/15/2025	15:47	lb135445
	Thallium	1050	1000	105	90 - 110	P	04/15/2025	15:47	lb135445
	Vanadium	474	500	95	90 - 110	P	04/15/2025	15:47	lb135445
	Zinc	1010	1000	100	90 - 110	P	04/15/2025	15:47	lb135445

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q1782

Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q1782 **SAS No.:** Q1782

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV	Antimony	51.5	50.0	103	80 - 120	P	04/15/2025	15:51	lb135445
	Arsenic	22.8	20.0	114	80 - 120	P	04/15/2025	15:51	lb135445
	Barium	95.1	100	95	80 - 120	P	04/15/2025	15:51	lb135445
	Beryllium	5.81	6.0	97	80 - 120	P	04/15/2025	15:51	lb135445
	Cadmium	5.96	6.0	99	80 - 120	P	04/15/2025	15:51	lb135445
	Chromium	9.28	10.0	93	80 - 120	P	04/15/2025	15:51	lb135445
	Cobalt	28.6	30.0	95	80 - 120	P	04/15/2025	15:51	lb135445
	Copper	21.4	20.0	107	80 - 120	P	04/15/2025	15:51	lb135445
	Iron	92.9	100	93	80 - 120	P	04/15/2025	15:51	lb135445
	Lead	12.1	12.0	101	80 - 120	P	04/15/2025	15:51	lb135445
	Manganese	20.5	20.0	103	80 - 120	P	04/15/2025	15:51	lb135445
	Nickel	38.4	40.0	96	80 - 120	P	04/15/2025	15:51	lb135445
	Potassium	2040	2000	102	80 - 120	P	04/15/2025	15:51	lb135445
	Selenium	22.5	20.0	113	80 - 120	P	04/15/2025	15:51	lb135445
	Silver	10.5	10.0	104	80 - 120	P	04/15/2025	15:51	lb135445
	Sodium	1820	2000	91	80 - 120	P	04/15/2025	15:51	lb135445
	Thallium	44.9	40.0	112	80 - 120	P	04/15/2025	15:51	lb135445
	Vanadium	38.8	40.0	97	80 - 120	P	04/15/2025	15:51	lb135445
	Zinc	42.3	40.0	106	80 - 120	P	04/15/2025	15:51	lb135445

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc. SDG No.: Q1782

Contract: LOCK01 Lab Code: CHEM Case No.: Q1782 SAS No.: Q1782

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Antimony	5020	5000	100	90 - 110	P	04/15/2025	16:26	lb135445
	Arsenic	5010	5000	100	90 - 110	P	04/15/2025	16:26	lb135445
	Barium	9820	10000	98	90 - 110	P	04/15/2025	16:26	lb135445
	Beryllium	249	250	100	90 - 110	P	04/15/2025	16:26	lb135445
	Cadmium	2440	2500	98	90 - 110	P	04/15/2025	16:26	lb135445
	Chromium	994	1000	99	90 - 110	P	04/15/2025	16:26	lb135445
	Cobalt	2450	2500	98	90 - 110	P	04/15/2025	16:26	lb135445
	Copper	1250	1250	100	90 - 110	P	04/15/2025	16:26	lb135445
	Iron	4940	5000	99	90 - 110	P	04/15/2025	16:26	lb135445
	Lead	4940	5000	99	90 - 110	P	04/15/2025	16:26	lb135445
	Manganese	2460	2500	98	90 - 110	P	04/15/2025	16:26	lb135445
	Nickel	2450	2500	98	90 - 110	P	04/15/2025	16:26	lb135445
	Potassium	25900	25000	104	90 - 110	P	04/15/2025	16:26	lb135445
	Selenium	5040	5000	101	90 - 110	P	04/15/2025	16:26	lb135445
	Silver	1240	1250	99	90 - 110	P	04/15/2025	16:26	lb135445
	Sodium	25600	25000	102	90 - 110	P	04/15/2025	16:26	lb135445
	Thallium	5080	5000	102	90 - 110	P	04/15/2025	16:26	lb135445
	Vanadium	2460	2500	98	90 - 110	P	04/15/2025	16:26	lb135445
	Zinc	2500	2500	100	90 - 110	P	04/15/2025	16:26	lb135445
CCV02	Antimony	4880	5000	98	90 - 110	P	04/15/2025	17:16	lb135445
	Arsenic	4870	5000	97	90 - 110	P	04/15/2025	17:16	lb135445
	Barium	9630	10000	96	90 - 110	P	04/15/2025	17:16	lb135445
	Beryllium	245	250	98	90 - 110	P	04/15/2025	17:16	lb135445
	Cadmium	2390	2500	96	90 - 110	P	04/15/2025	17:16	lb135445
	Chromium	988	1000	99	90 - 110	P	04/15/2025	17:16	lb135445
	Cobalt	2400	2500	96	90 - 110	P	04/15/2025	17:16	lb135445
	Copper	1220	1250	98	90 - 110	P	04/15/2025	17:16	lb135445
	Iron	4870	5000	97	90 - 110	P	04/15/2025	17:16	lb135445
	Lead	4820	5000	96	90 - 110	P	04/15/2025	17:16	lb135445
	Manganese	2410	2500	96	90 - 110	P	04/15/2025	17:16	lb135445
	Nickel	2400	2500	96	90 - 110	P	04/15/2025	17:16	lb135445
	Potassium	25400	25000	102	90 - 110	P	04/15/2025	17:16	lb135445
	Selenium	4880	5000	98	90 - 110	P	04/15/2025	17:16	lb135445
	Silver	1230	1250	98	90 - 110	P	04/15/2025	17:16	lb135445

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc. SDG No.: Q1782

Contract: LOCK01 Lab Code: CHEM Case No.: Q1782 SAS No.: Q1782

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Sodium	25100	25000	100	90 - 110	P	04/15/2025	17:16	lb135445
	Thallium	4960	5000	99	90 - 110	P	04/15/2025	17:16	lb135445
	Vanadium	2420	2500	97	90 - 110	P	04/15/2025	17:16	lb135445
	Zinc	2460	2500	98	90 - 110	P	04/15/2025	17:16	lb135445
CCV03	Antimony	4980	5000	100	90 - 110	P	04/15/2025	18:07	lb135445
	Arsenic	4970	5000	99	90 - 110	P	04/15/2025	18:07	lb135445
	Barium	9800	10000	98	90 - 110	P	04/15/2025	18:07	lb135445
	Beryllium	242	250	97	90 - 110	P	04/15/2025	18:07	lb135445
	Cadmium	2440	2500	98	90 - 110	P	04/15/2025	18:07	lb135445
	Chromium	996	1000	100	90 - 110	P	04/15/2025	18:07	lb135445
	Cobalt	2440	2500	98	90 - 110	P	04/15/2025	18:07	lb135445
	Copper	1240	1250	99	90 - 110	P	04/15/2025	18:07	lb135445
	Iron	5050	5000	101	90 - 110	P	04/15/2025	18:07	lb135445
	Lead	4910	5000	98	90 - 110	P	04/15/2025	18:07	lb135445
	Manganese	2420	2500	97	90 - 110	P	04/15/2025	18:07	lb135445
	Nickel	2440	2500	98	90 - 110	P	04/15/2025	18:07	lb135445
	Potassium	25200	25000	101	90 - 110	P	04/15/2025	18:07	lb135445
	Selenium	5010	5000	100	90 - 110	P	04/15/2025	18:07	lb135445
	Silver	1240	1250	99	90 - 110	P	04/15/2025	18:07	lb135445
	Sodium	25000	25000	100	90 - 110	P	04/15/2025	18:07	lb135445
	Thallium	5020	5000	100	90 - 110	P	04/15/2025	18:07	lb135445
	Vanadium	2440	2500	98	90 - 110	P	04/15/2025	18:07	lb135445
	Zinc	2490	2500	100	90 - 110	P	04/15/2025	18:07	lb135445
CCV04	Antimony	4970	5000	100	90 - 110	P	04/15/2025	19:08	lb135445
	Arsenic	4980	5000	100	90 - 110	P	04/15/2025	19:08	lb135445
	Barium	9680	10000	97	90 - 110	P	04/15/2025	19:08	lb135445
	Beryllium	246	250	98	90 - 110	P	04/15/2025	19:08	lb135445
	Cadmium	2460	2500	98	90 - 110	P	04/15/2025	19:08	lb135445
	Chromium	994	1000	99	90 - 110	P	04/15/2025	19:08	lb135445
	Cobalt	2460	2500	98	90 - 110	P	04/15/2025	19:08	lb135445
	Copper	1240	1250	99	90 - 110	P	04/15/2025	19:08	lb135445
	Iron	4910	5000	98	90 - 110	P	04/15/2025	19:08	lb135445
	Lead	4950	5000	99	90 - 110	P	04/15/2025	19:08	lb135445
	Manganese	2440	2500	97	90 - 110	P	04/15/2025	19:08	lb135445

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc. SDG No.: Q1782
 Contract: LOCK01 Lab Code: CHEM Case No.: Q1782 SAS No.: Q1782
 Initial Calibration Source: EPA
 Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Nickel	2460	2500	98	90 - 110	P	04/15/2025	19:08	lb135445
	Potassium	24400	25000	98	90 - 110	P	04/15/2025	19:08	lb135445
	Selenium	5000	5000	100	90 - 110	P	04/15/2025	19:08	lb135445
	Silver	1230	1250	98	90 - 110	P	04/15/2025	19:08	lb135445
	Sodium	24100	25000	96	90 - 110	P	04/15/2025	19:08	lb135445
	Thallium	5020	5000	100	90 - 110	P	04/15/2025	19:08	lb135445
	Vanadium	2460	2500	98	90 - 110	P	04/15/2025	19:08	lb135445
	Zinc	2470	2500	99	90 - 110	P	04/15/2025	19:08	lb135445
CCV05	Antimony	4830	5000	96	90 - 110	P	04/15/2025	19:59	lb135445
	Arsenic	4820	5000	96	90 - 110	P	04/15/2025	19:59	lb135445
	Barium	9600	10000	96	90 - 110	P	04/15/2025	19:59	lb135445
	Beryllium	238	250	95	90 - 110	P	04/15/2025	19:59	lb135445
	Cadmium	2390	2500	96	90 - 110	P	04/15/2025	19:59	lb135445
	Chromium	972	1000	97	90 - 110	P	04/15/2025	19:59	lb135445
	Cobalt	2390	2500	96	90 - 110	P	04/15/2025	19:59	lb135445
	Copper	1200	1250	96	90 - 110	P	04/15/2025	19:59	lb135445
	Iron	4870	5000	97	90 - 110	P	04/15/2025	19:59	lb135445
	Lead	4830	5000	97	90 - 110	P	04/15/2025	19:59	lb135445
	Manganese	2390	2500	96	90 - 110	P	04/15/2025	19:59	lb135445
	Nickel	2400	2500	96	90 - 110	P	04/15/2025	19:59	lb135445
	Potassium	24100	25000	97	90 - 110	P	04/15/2025	19:59	lb135445
	Selenium	4840	5000	97	90 - 110	P	04/15/2025	19:59	lb135445
	Silver	1210	1250	96	90 - 110	P	04/15/2025	19:59	lb135445
	Sodium	24000	25000	96	90 - 110	P	04/15/2025	19:59	lb135445
	Thallium	4850	5000	97	90 - 110	P	04/15/2025	19:59	lb135445
	Vanadium	2400	2500	96	90 - 110	P	04/15/2025	19:59	lb135445
	Zinc	2410	2500	96	90 - 110	P	04/15/2025	19:59	lb135445
CCV06	Antimony	4980	5000	100	90 - 110	P	04/15/2025	20:51	lb135445
	Arsenic	4980	5000	100	90 - 110	P	04/15/2025	20:51	lb135445
	Barium	9450	10000	94	90 - 110	P	04/15/2025	20:51	lb135445
	Beryllium	245	250	98	90 - 110	P	04/15/2025	20:51	lb135445
	Cadmium	2460	2500	98	90 - 110	P	04/15/2025	20:51	lb135445
	Chromium	1010	1000	100	90 - 110	P	04/15/2025	20:51	lb135445
	Cobalt	2450	2500	98	90 - 110	P	04/15/2025	20:51	lb135445

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q1782

Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q1782 **SAS No.:** Q1782

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV06	Copper	1230	1250	98	90 - 110	P	04/15/2025	20:51	lb135445
	Iron	4940	5000	99	90 - 110	P	04/15/2025	20:51	lb135445
	Lead	4960	5000	99	90 - 110	P	04/15/2025	20:51	lb135445
	Manganese	2380	2500	95	90 - 110	P	04/15/2025	20:51	lb135445
	Nickel	2470	2500	99	90 - 110	P	04/15/2025	20:51	lb135445
	Potassium	25300	25000	101	90 - 110	P	04/15/2025	20:51	lb135445
	Selenium	4980	5000	100	90 - 110	P	04/15/2025	20:51	lb135445
	Silver	1240	1250	99	90 - 110	P	04/15/2025	20:51	lb135445
	Sodium	24900	25000	100	90 - 110	P	04/15/2025	20:51	lb135445
	Thallium	5000	5000	100	90 - 110	P	04/15/2025	20:51	lb135445
	Vanadium	2430	2500	97	90 - 110	P	04/15/2025	20:51	lb135445
	Zinc	2490	2500	99	90 - 110	P	04/15/2025	20:51	lb135445
CCV07	Antimony	4940	5000	99	90 - 110	P	04/15/2025	21:43	lb135445
	Arsenic	4960	5000	99	90 - 110	P	04/15/2025	21:43	lb135445
	Barium	9520	10000	95	90 - 110	P	04/15/2025	21:43	lb135445
	Beryllium	246	250	98	90 - 110	P	04/15/2025	21:43	lb135445
	Cadmium	2460	2500	98	90 - 110	P	04/15/2025	21:43	lb135445
	Chromium	991	1000	99	90 - 110	P	04/15/2025	21:43	lb135445
	Cobalt	2450	2500	98	90 - 110	P	04/15/2025	21:43	lb135445
	Copper	1220	1250	98	90 - 110	P	04/15/2025	21:43	lb135445
	Iron	4840	5000	97	90 - 110	P	04/15/2025	21:43	lb135445
	Lead	4950	5000	99	90 - 110	P	04/15/2025	21:43	lb135445
	Manganese	2370	2500	95	90 - 110	P	04/15/2025	21:43	lb135445
	Nickel	2480	2500	99	90 - 110	P	04/15/2025	21:43	lb135445
	Potassium	24200	25000	97	90 - 110	P	04/15/2025	21:43	lb135445
	Selenium	5000	5000	100	90 - 110	P	04/15/2025	21:43	lb135445
	Silver	1230	1250	98	90 - 110	P	04/15/2025	21:43	lb135445
	Sodium	23900	25000	96	90 - 110	P	04/15/2025	21:43	lb135445
	Thallium	4850	5000	97	90 - 110	P	04/15/2025	21:43	lb135445
	Vanadium	2410	2500	96	90 - 110	P	04/15/2025	21:43	lb135445
	Zinc	2460	2500	98	90 - 110	P	04/15/2025	21:43	lb135445
	Antimony	4900	5000	98	90 - 110	P	04/15/2025	22:25	lb135445
CCV08	Arsenic	4920	5000	98	90 - 110	P	04/15/2025	22:25	lb135445
	Barium	9440	10000	94	90 - 110	P	04/15/2025	22:25	lb135445

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Lockwood, Kessler & Bartlett, Inc. SDG No.: Q1782
 Contract: LOCK01 Lab Code: CHEM Case No.: Q1782 SAS No.: Q1782
 Initial Calibration Source: EPA
 Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Beryllium	248	250	99	90 - 110	P	04/15/2025	22:25	lb135445
	Cadmium	2460	2500	98	90 - 110	P	04/15/2025	22:25	lb135445
	Chromium	993	1000	99	90 - 110	P	04/15/2025	22:25	lb135445
	Cobalt	2440	2500	98	90 - 110	P	04/15/2025	22:25	lb135445
	Copper	1210	1250	97	90 - 110	P	04/15/2025	22:25	lb135445
	Iron	4820	5000	96	90 - 110	P	04/15/2025	22:25	lb135445
	Lead	4940	5000	99	90 - 110	P	04/15/2025	22:25	lb135445
	Manganese	2370	2500	95	90 - 110	P	04/15/2025	22:25	lb135445
	Nickel	2470	2500	99	90 - 110	P	04/15/2025	22:25	lb135445
	Potassium	24000	25000	96	90 - 110	P	04/15/2025	22:25	lb135445
	Selenium	4940	5000	99	90 - 110	P	04/15/2025	22:25	lb135445
	Silver	1230	1250	98	90 - 110	P	04/15/2025	22:25	lb135445
	Sodium	23500	25000	94	90 - 110	P	04/15/2025	22:25	lb135445
	Thallium	4950	5000	99	90 - 110	P	04/15/2025	22:25	lb135445
	Vanadium	2410	2500	96	90 - 110	P	04/15/2025	22:25	lb135445
	Zinc	2460	2500	98	90 - 110	P	04/15/2025	22:25	lb135445



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q1782
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q1782 **SAS No.:** Q1782
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI1	Antimony	54.2	50.0	108	40 - 160	P	04/15/2025	16:05	lb135445
	Arsenic	23.2	20.0	116	40 - 160	P	04/15/2025	16:05	lb135445
	Barium	93.9	100	94	40 - 160	P	04/15/2025	16:05	lb135445
	Beryllium	5.90	6.0	98	40 - 160	P	04/15/2025	16:05	lb135445
	Cadmium	5.84	6.0	97	40 - 160	P	04/15/2025	16:05	lb135445
	Chromium	9.65	10.0	97	40 - 160	P	04/15/2025	16:05	lb135445
	Cobalt	28.6	30.0	95	40 - 160	P	04/15/2025	16:05	lb135445
	Copper	21.4	20.0	107	40 - 160	P	04/15/2025	16:05	lb135445
	Iron	90.6	100	91	40 - 160	P	04/15/2025	16:05	lb135445
	Lead	12.5	12.0	104	40 - 160	P	04/15/2025	16:05	lb135445
	Manganese	20.9	20.0	104	40 - 160	P	04/15/2025	16:05	lb135445
	Nickel	38.5	40.0	96	40 - 160	P	04/15/2025	16:05	lb135445
	Potassium	2000	2000	100	40 - 160	P	04/15/2025	16:05	lb135445
	Selenium	22.6	20.0	113	40 - 160	P	04/15/2025	16:05	lb135445
	Silver	10.3	10.0	103	40 - 160	P	04/15/2025	16:05	lb135445
	Sodium	1770	2000	88	40 - 160	P	04/15/2025	16:05	lb135445
	Thallium	43.2	40.0	108	40 - 160	P	04/15/2025	16:05	lb135445
	Vanadium	39.3	40.0	98	40 - 160	P	04/15/2025	16:05	lb135445
	Zinc	41.4	40.0	104	40 - 160	P	04/15/2025	16:05	lb135445



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q1782
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q1782 **SAS No.:** Q1782

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Antimony	50.0	+/-50.0	U	50.0	P	04/15/2025	15:55	lb135445
	Arsenic	20.0	+/-20.0	U	20.0	P	04/15/2025	15:55	lb135445
	Barium	100	+/-100	U	100	P	04/15/2025	15:55	lb135445
	Beryllium	6.00	+/-6.00	U	6.00	P	04/15/2025	15:55	lb135445
	Cadmium	6.00	+/-6.00	U	6.00	P	04/15/2025	15:55	lb135445
	Chromium	10.0	+/-10.0	U	10.0	P	04/15/2025	15:55	lb135445
	Cobalt	30.0	+/-30.0	U	30.0	P	04/15/2025	15:55	lb135445
	Copper	20.0	+/-20.0	U	20.0	P	04/15/2025	15:55	lb135445
	Iron	100	+/-100	U	100	P	04/15/2025	15:55	lb135445
	Lead	12.0	+/-12.0	U	12.0	P	04/15/2025	15:55	lb135445
	Manganese	20.0	+/-20.0	U	20.0	P	04/15/2025	15:55	lb135445
	Nickel	40.0	+/-40.0	U	40.0	P	04/15/2025	15:55	lb135445
	Potassium	2000	+/-2000	U	2000	P	04/15/2025	15:55	lb135445
	Selenium	20.0	+/-20.0	U	20.0	P	04/15/2025	15:55	lb135445
	Silver	10.0	+/-10.0	U	10.0	P	04/15/2025	15:55	lb135445
	Sodium	2000	+/-2000	U	2000	P	04/15/2025	15:55	lb135445
	Thallium	40.0	+/-40.0	U	40.0	P	04/15/2025	15:55	lb135445
	Vanadium	40.0	+/-40.0	U	40.0	P	04/15/2025	15:55	lb135445
	Zinc	40.0	+/-40.0	U	40.0	P	04/15/2025	15:55	lb135445

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Lockwood, Kessler & Bartlett, Inc.				SDG No.:	Q1782			
Contract:	LOCK01	Lab Code:	CHEM	Case No.:	Q1782	SAS No.:	Q1782		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Antimony	50.0	+/-50.0	U	50.0	P	04/15/2025	16:30	lb135445
	Arsenic	20.0	+/-20.0	U	20.0	P	04/15/2025	16:30	lb135445
	Barium	100	+/-100	U	100	P	04/15/2025	16:30	lb135445
	Beryllium	6.00	+/-6.00	U	6.00	P	04/15/2025	16:30	lb135445
	Cadmium	6.00	+/-6.00	U	6.00	P	04/15/2025	16:30	lb135445
	Chromium	10.0	+/-10.0	U	10.0	P	04/15/2025	16:30	lb135445
	Cobalt	30.0	+/-30.0	U	30.0	P	04/15/2025	16:30	lb135445
	Copper	20.0	+/-20.0	U	20.0	P	04/15/2025	16:30	lb135445
	Iron	100	+/-100	U	100	P	04/15/2025	16:30	lb135445
	Lead	12.0	+/-12.0	U	12.0	P	04/15/2025	16:30	lb135445
	Manganese	20.0	+/-20.0	U	20.0	P	04/15/2025	16:30	lb135445
	Nickel	40.0	+/-40.0	U	40.0	P	04/15/2025	16:30	lb135445
	Potassium	2000	+/-2000	U	2000	P	04/15/2025	16:30	lb135445
	Selenium	20.0	+/-20.0	U	20.0	P	04/15/2025	16:30	lb135445
	Silver	10.0	+/-10.0	U	10.0	P	04/15/2025	16:30	lb135445
	Sodium	2000	+/-2000	U	2000	P	04/15/2025	16:30	lb135445
	Thallium	40.0	+/-40.0	U	40.0	P	04/15/2025	16:30	lb135445
	Vanadium	40.0	+/-40.0	U	40.0	P	04/15/2025	16:30	lb135445
	Zinc	40.0	+/-40.0	U	40.0	P	04/15/2025	16:30	lb135445
CCB02	Antimony	50.0	+/-50.0	U	50.0	P	04/15/2025	17:20	lb135445
	Arsenic	20.0	+/-20.0	U	20.0	P	04/15/2025	17:20	lb135445
	Barium	100	+/-100	U	100	P	04/15/2025	17:20	lb135445
	Beryllium	6.00	+/-6.00	U	6.00	P	04/15/2025	17:20	lb135445
	Cadmium	6.00	+/-6.00	U	6.00	P	04/15/2025	17:20	lb135445
	Chromium	10.0	+/-10.0	U	10.0	P	04/15/2025	17:20	lb135445
	Cobalt	30.0	+/-30.0	U	30.0	P	04/15/2025	17:20	lb135445
	Copper	20.0	+/-20.0	U	20.0	P	04/15/2025	17:20	lb135445
	Iron	100	+/-100	U	100	P	04/15/2025	17:20	lb135445
	Lead	12.0	+/-12.0	U	12.0	P	04/15/2025	17:20	lb135445
	Manganese	20.0	+/-20.0	U	20.0	P	04/15/2025	17:20	lb135445
	Nickel	40.0	+/-40.0	U	40.0	P	04/15/2025	17:20	lb135445
	Potassium	2000	+/-2000	U	2000	P	04/15/2025	17:20	lb135445
	Selenium	20.0	+/-20.0	U	20.0	P	04/15/2025	17:20	lb135445
	Silver	10.0	+/-10.0	U	10.0	P	04/15/2025	17:20	lb135445
	Sodium	2000	+/-2000	U	2000	P	04/15/2025	17:20	lb135445
	Thallium	40.0	+/-40.0	U	40.0	P	04/15/2025	17:20	lb135445
	Vanadium	40.0	+/-40.0	U	40.0	P	04/15/2025	17:20	lb135445
	Zinc	40.0	+/-40.0	U	40.0	P	04/15/2025	17:20	lb135445
CCB03	Antimony	50.0	+/-50.0	U	50.0	P	04/15/2025	18:11	lb135445

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Lockwood, Kessler & Bartlett, Inc.</u>		SDG No.: <u>Q1782</u>							
Contract: <u>LOCK01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1782</u>	SAS No.: <u>Q1782</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	04/15/2025	18:11	lb135445
	Barium	100	+/-100	U	100	P	04/15/2025	18:11	lb135445
	Beryllium	6.00	+/-6.00	U	6.00	P	04/15/2025	18:11	lb135445
	Cadmium	6.00	+/-6.00	U	6.00	P	04/15/2025	18:11	lb135445
	Chromium	10.0	+/-10.0	U	10.0	P	04/15/2025	18:11	lb135445
	Cobalt	30.0	+/-30.0	U	30.0	P	04/15/2025	18:11	lb135445
	Copper	20.0	+/-20.0	U	20.0	P	04/15/2025	18:11	lb135445
	Iron	100	+/-100	U	100	P	04/15/2025	18:11	lb135445
	Lead	12.0	+/-12.0	U	12.0	P	04/15/2025	18:11	lb135445
	Manganese	20.0	+/-20.0	U	20.0	P	04/15/2025	18:11	lb135445
	Nickel	40.0	+/-40.0	U	40.0	P	04/15/2025	18:11	lb135445
	Potassium	2000	+/-2000	U	2000	P	04/15/2025	18:11	lb135445
	Selenium	20.0	+/-20.0	U	20.0	P	04/15/2025	18:11	lb135445
	Silver	10.0	+/-10.0	U	10.0	P	04/15/2025	18:11	lb135445
	Sodium	2000	+/-2000	U	2000	P	04/15/2025	18:11	lb135445
	Thallium	40.0	+/-40.0	U	40.0	P	04/15/2025	18:11	lb135445
	Vanadium	40.0	+/-40.0	U	40.0	P	04/15/2025	18:11	lb135445
	Zinc	40.0	+/-40.0	U	40.0	P	04/15/2025	18:11	lb135445
CCB04	Antimony	50.0	+/-50.0	U	50.0	P	04/15/2025	19:12	lb135445
	Arsenic	20.0	+/-20.0	U	20.0	P	04/15/2025	19:12	lb135445
	Barium	100	+/-100	U	100	P	04/15/2025	19:12	lb135445
	Beryllium	6.00	+/-6.00	U	6.00	P	04/15/2025	19:12	lb135445
	Cadmium	6.00	+/-6.00	U	6.00	P	04/15/2025	19:12	lb135445
	Chromium	10.0	+/-10.0	U	10.0	P	04/15/2025	19:12	lb135445
	Cobalt	30.0	+/-30.0	U	30.0	P	04/15/2025	19:12	lb135445
	Copper	20.0	+/-20.0	U	20.0	P	04/15/2025	19:12	lb135445
	Iron	100	+/-100	U	100	P	04/15/2025	19:12	lb135445
	Lead	12.0	+/-12.0	U	12.0	P	04/15/2025	19:12	lb135445
	Manganese	20.0	+/-20.0	U	20.0	P	04/15/2025	19:12	lb135445
	Nickel	40.0	+/-40.0	U	40.0	P	04/15/2025	19:12	lb135445
	Potassium	2000	+/-2000	U	2000	P	04/15/2025	19:12	lb135445
	Selenium	20.0	+/-20.0	U	20.0	P	04/15/2025	19:12	lb135445
	Silver	10.0	+/-10.0	U	10.0	P	04/15/2025	19:12	lb135445
	Sodium	2000	+/-2000	U	2000	P	04/15/2025	19:12	lb135445
	Thallium	40.0	+/-40.0	U	40.0	P	04/15/2025	19:12	lb135445
	Vanadium	40.0	+/-40.0	U	40.0	P	04/15/2025	19:12	lb135445
	Zinc	40.0	+/-40.0	U	40.0	P	04/15/2025	19:12	lb135445
CCB05	Antimony	50.0	+/-50.0	U	50.0	P	04/15/2025	20:03	lb135445
	Arsenic	20.0	+/-20.0	U	20.0	P	04/15/2025	20:03	lb135445

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.		SDG No.: Q1782							
Contract: LOCK01	Lab Code: CHEM	Case No.: Q1782	SAS No.: Q1782						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB05	Barium	100	+/-100	U	100	P	04/15/2025	20:03	lb135445
	Beryllium	6.00	+/-6.00	U	6.00	P	04/15/2025	20:03	lb135445
	Cadmium	6.00	+/-6.00	U	6.00	P	04/15/2025	20:03	lb135445
	Chromium	10.0	+/-10.0	U	10.0	P	04/15/2025	20:03	lb135445
	Cobalt	30.0	+/-30.0	U	30.0	P	04/15/2025	20:03	lb135445
	Copper	20.0	+/-20.0	U	20.0	P	04/15/2025	20:03	lb135445
	Iron	100	+/-100	U	100	P	04/15/2025	20:03	lb135445
	Lead	12.0	+/-12.0	U	12.0	P	04/15/2025	20:03	lb135445
	Manganese	20.0	+/-20.0	U	20.0	P	04/15/2025	20:03	lb135445
	Nickel	40.0	+/-40.0	U	40.0	P	04/15/2025	20:03	lb135445
	Potassium	2000	+/-2000	U	2000	P	04/15/2025	20:03	lb135445
	Selenium	20.0	+/-20.0	U	20.0	P	04/15/2025	20:03	lb135445
	Silver	10.0	+/-10.0	U	10.0	P	04/15/2025	20:03	lb135445
	Sodium	2000	+/-2000	U	2000	P	04/15/2025	20:03	lb135445
	Thallium	40.0	+/-40.0	U	40.0	P	04/15/2025	20:03	lb135445
	Vanadium	40.0	+/-40.0	U	40.0	P	04/15/2025	20:03	lb135445
	Zinc	40.0	+/-40.0	U	40.0	P	04/15/2025	20:03	lb135445
CCB06	Antimony	50.0	+/-50.0	U	50.0	P	04/15/2025	20:55	lb135445
	Arsenic	20.0	+/-20.0	U	20.0	P	04/15/2025	20:55	lb135445
	Barium	100	+/-100	U	100	P	04/15/2025	20:55	lb135445
	Beryllium	6.00	+/-6.00	U	6.00	P	04/15/2025	20:55	lb135445
	Cadmium	6.00	+/-6.00	U	6.00	P	04/15/2025	20:55	lb135445
	Chromium	10.0	+/-10.0	U	10.0	P	04/15/2025	20:55	lb135445
	Cobalt	30.0	+/-30.0	U	30.0	P	04/15/2025	20:55	lb135445
	Copper	20.0	+/-20.0	U	20.0	P	04/15/2025	20:55	lb135445
	Iron	100	+/-100	U	100	P	04/15/2025	20:55	lb135445
	Lead	12.0	+/-12.0	U	12.0	P	04/15/2025	20:55	lb135445
	Manganese	20.0	+/-20.0	U	20.0	P	04/15/2025	20:55	lb135445
	Nickel	40.0	+/-40.0	U	40.0	P	04/15/2025	20:55	lb135445
	Potassium	2000	+/-2000	U	2000	P	04/15/2025	20:55	lb135445
	Selenium	20.0	+/-20.0	U	20.0	P	04/15/2025	20:55	lb135445
	Silver	10.0	+/-10.0	U	10.0	P	04/15/2025	20:55	lb135445
	Sodium	2000	+/-2000	U	2000	P	04/15/2025	20:55	lb135445
	Thallium	40.0	+/-40.0	U	40.0	P	04/15/2025	20:55	lb135445
	Vanadium	40.0	+/-40.0	U	40.0	P	04/15/2025	20:55	lb135445
	Zinc	40.0	+/-40.0	U	40.0	P	04/15/2025	20:55	lb135445
CCB07	Antimony	50.0	+/-50.0	U	50.0	P	04/15/2025	21:47	lb135445
	Arsenic	20.0	+/-20.0	U	20.0	P	04/15/2025	21:47	lb135445
	Barium	100	+/-100	U	100	P	04/15/2025	21:47	lb135445

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q1782
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q1782 **SAS No.:** Q1782

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB07	Beryllium	6.00	+/-6.00	U	6.00	P	04/15/2025	21:47	lb135445
	Cadmium	6.00	+/-6.00	U	6.00	P	04/15/2025	21:47	lb135445
	Chromium	10.0	+/-10.0	U	10.0	P	04/15/2025	21:47	lb135445
	Cobalt	30.0	+/-30.0	U	30.0	P	04/15/2025	21:47	lb135445
	Copper	20.0	+/-20.0	U	20.0	P	04/15/2025	21:47	lb135445
	Iron	100	+/-100	U	100	P	04/15/2025	21:47	lb135445
	Lead	12.0	+/-12.0	U	12.0	P	04/15/2025	21:47	lb135445
	Manganese	20.0	+/-20.0	U	20.0	P	04/15/2025	21:47	lb135445
	Nickel	40.0	+/-40.0	U	40.0	P	04/15/2025	21:47	lb135445
	Potassium	2000	+/-2000	U	2000	P	04/15/2025	21:47	lb135445
	Selenium	20.0	+/-20.0	U	20.0	P	04/15/2025	21:47	lb135445
	Silver	10.0	+/-10.0	U	10.0	P	04/15/2025	21:47	lb135445
	Sodium	2000	+/-2000	U	2000	P	04/15/2025	21:47	lb135445
	Thallium	40.0	+/-40.0	U	40.0	P	04/15/2025	21:47	lb135445
	Vanadium	40.0	+/-40.0	U	40.0	P	04/15/2025	21:47	lb135445
	Zinc	40.0	+/-40.0	U	40.0	P	04/15/2025	21:47	lb135445
CCB08	Antimony	50.0	+/-50.0	U	50.0	P	04/15/2025	22:29	lb135445
	Arsenic	20.0	+/-20.0	U	20.0	P	04/15/2025	22:29	lb135445
	Barium	100	+/-100	U	100	P	04/15/2025	22:29	lb135445
	Beryllium	6.00	+/-6.00	U	6.00	P	04/15/2025	22:29	lb135445
	Cadmium	6.00	+/-6.00	U	6.00	P	04/15/2025	22:29	lb135445
	Chromium	10.0	+/-10.0	U	10.0	P	04/15/2025	22:29	lb135445
	Cobalt	30.0	+/-30.0	U	30.0	P	04/15/2025	22:29	lb135445
	Copper	20.0	+/-20.0	U	20.0	P	04/15/2025	22:29	lb135445
	Iron	100	+/-100	U	100	P	04/15/2025	22:29	lb135445
	Lead	12.0	+/-12.0	U	12.0	P	04/15/2025	22:29	lb135445
	Manganese	20.0	+/-20.0	U	20.0	P	04/15/2025	22:29	lb135445
	Nickel	40.0	+/-40.0	U	40.0	P	04/15/2025	22:29	lb135445
	Potassium	2000	+/-2000	U	2000	P	04/15/2025	22:29	lb135445
	Selenium	20.0	+/-20.0	U	20.0	P	04/15/2025	22:29	lb135445
	Silver	10.0	+/-10.0	U	10.0	P	04/15/2025	22:29	lb135445
	Sodium	2000	+/-2000	U	2000	P	04/15/2025	22:29	lb135445
	Thallium	40.0	+/-40.0	U	40.0	P	04/15/2025	22:29	lb135445
	Vanadium	40.0	+/-40.0	U	40.0	P	04/15/2025	22:29	lb135445
	Zinc	40.0	+/-40.0	U	40.0	P	04/15/2025	22:29	lb135445

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q1782

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB167583BL	WATER			Batch Number:	PB167583		Prep Date:	04/14/2025	
	Antimony	25.0	<25.0	U	25.0	P	04/15/2025	19:33	lb135445
	Arsenic	10.0	<10.0	U	10.0	P	04/15/2025	19:33	lb135445
	Barium	50.0	<50.0	U	50.0	P	04/15/2025	19:33	lb135445
	Beryllium	3.00	<3.00	U	3.00	P	04/15/2025	19:33	lb135445
	Cadmium	3.00	<3.00	U	3.00	P	04/15/2025	19:33	lb135445
	Chromium	5.00	<5.00	U	5.00	P	04/15/2025	19:33	lb135445
	Cobalt	15.0	<15.0	U	15.0	P	04/15/2025	19:33	lb135445
	Copper	10.0	<10.0	U	10.0	P	04/15/2025	19:33	lb135445
	Iron	50.0	<50.0	U	50.0	P	04/15/2025	19:33	lb135445
	Lead	6.00	<6.00	U	6.00	P	04/15/2025	19:33	lb135445
	Manganese	10.0	<10.0	U	10.0	P	04/15/2025	19:33	lb135445
	Nickel	20.0	<20.0	U	20.0	P	04/15/2025	19:33	lb135445
	Potassium	1000	<1000	U	1000	P	04/15/2025	19:33	lb135445
	Selenium	10.0	<10.0	U	10.0	P	04/15/2025	19:33	lb135445
	Silver	5.00	<5.00	U	5.00	P	04/15/2025	19:33	lb135445
	Sodium	1000	<1000	U	1000	P	04/15/2025	19:33	lb135445
	Thallium	20.0	<20.0	U	20.0	P	04/15/2025	19:33	lb135445
	Vanadium	20.0	<20.0	U	20.0	P	04/15/2025	19:33	lb135445
	Zinc	20.0	<20.0	U	20.0	P	04/15/2025	19:33	lb135445

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q1782
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q1782 **SAS No.:** Q1782
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Antimony	11.2			-50	50	04/15/2025	16:09	lb135445
	Arsenic	1.49			-20	20	04/15/2025	16:09	lb135445
	Barium	32.8	6.0	547	-94	106	04/15/2025	16:09	lb135445
	Beryllium	1.40			-6	6	04/15/2025	16:09	lb135445
	Cadmium	4.88	1.0	488	-5	7	04/15/2025	16:09	lb135445
	Chromium	57.8	52.0	111	42	62	04/15/2025	16:09	lb135445
	Cobalt	9.75			-30	30	04/15/2025	16:09	lb135445
	Copper	15.2	2.0	760	-18	22	04/15/2025	16:09	lb135445
	Iron	96800	101000	96	85600	116500	04/15/2025	16:09	lb135445
	Lead	11.2			-12	12	04/15/2025	16:09	lb135445
	Manganese	15.7	7.0	224	-13	27	04/15/2025	16:09	lb135445
	Nickel	9.59	2.0	480	-38	42	04/15/2025	16:09	lb135445
	Potassium	172			0	0	04/15/2025	16:09	lb135445
	Selenium	10.3			-20	20	04/15/2025	16:09	lb135445
	Silver	-2.32			-10	10	04/15/2025	16:09	lb135445
	Sodium	96.5			0	0	04/15/2025	16:09	lb135445
	Thallium	25.3			-40	40	04/15/2025	16:09	lb135445
	Vanadium	10.1			-40	40	04/15/2025	16:09	lb135445
	Zinc	11.8			-40	40	04/15/2025	16:09	lb135445
ICSAB01	Antimony	643	618	104	525	711	04/15/2025	16:13	lb135445
	Arsenic	96.5	104	93	88.4	120	04/15/2025	16:13	lb135445
	Barium	467	537	87	437	637	04/15/2025	16:13	lb135445
	Beryllium	490	495	99	420	570	04/15/2025	16:13	lb135445
	Cadmium	1020	972	105	826	1120	04/15/2025	16:13	lb135445
	Chromium	561	542	104	460	624	04/15/2025	16:13	lb135445
	Cobalt	497	476	104	404	548	04/15/2025	16:13	lb135445
	Copper	498	511	98	434	588	04/15/2025	16:13	lb135445
	Iron	94700	99300	95	84400	114500	04/15/2025	16:13	lb135445
	Lead	41.0	49.0	84	37	61	04/15/2025	16:13	lb135445
	Manganese	471	507	93	430	584	04/15/2025	16:13	lb135445
	Nickel	980	954	103	810	1100	04/15/2025	16:13	lb135445
	Potassium	53.4			0	0	04/15/2025	16:13	lb135445
	Selenium	49.4	46.0	107	26	66	04/15/2025	16:13	lb135445
	Silver	177	201	88	170	232	04/15/2025	16:13	lb135445
	Sodium	-14.9			0	0	04/15/2025	16:13	lb135445
	Thallium	101	108	94	68	148	04/15/2025	16:13	lb135445
	Vanadium	467	491	95	417	565	04/15/2025	16:13	lb135445
	Zinc	1040	952	109	809	1095	04/15/2025	16:13	lb135445



METAL QC DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client:	<u>Lockwood, Kessler & Bartlett, Inc.</u>	level:	<u>low</u>	sdg no.:	<u>Q1782</u>
contract:	<u>LOCK01</u>	lab code:	<u>CHEM</u>	case no.:	<u>Q1782</u>
matrix:	<u>Water</u>	sample id:	<u>Q1782-07</u>	client id:	<u>MW-4MS</u>
Percent Solids for Sample:	<u>NA</u>	Spiked ID:	<u>Q1782-07MS</u>	Percent Solids for Spike Sample:	<u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Antimony	ug/L	75 - 125	350		25.0	U	400	88		P
Arsenic	ug/L	75 - 125	347		10.0	U	400	87		P
Barium	ug/L	75 - 125	191		113		100	78		P
Beryllium	ug/L	75 - 125	77.8		3.00	U	100	78		P
Cadmium	ug/L	75 - 125	84.7		3.00	U	100	85		P
Chromium	ug/L	75 - 125	163		5.00	U	200	82		P
Cobalt	ug/L	75 - 125	81.6		15.0	U	100	82		P
Copper	ug/L	75 - 125	119		10.0	U	150	79		P
Iron	ug/L	75 - 125	1330		67.2		1500	84		P
Lead	ug/L	75 - 125	404		6.00	U	500	81		P
Manganese	ug/L	75 - 125	106		98.1		100	8	N	P
Nickel	ug/L	75 - 125	204		20.0	U	250	82		P
Potassium	ug/L	75 - 125	16400		12200		5000	85		P
Selenium	ug/L	75 - 125	832		10.0	U	1000	83		P
Silver	ug/L	75 - 125	31.2		5.00	U	37.5	83		P
Sodium	ug/L	75 - 125	198000		202000		1500	-221		P
Thallium	ug/L	75 - 125	780		20.0	U	1000	78		P
Vanadium	ug/L	75 - 125	119		20.0	U	150	79		P
Zinc	ug/L	75 - 125	93.8		9.56	J	100	84		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client:	<u>Lockwood, Kessler & Bartlett, Inc.</u>	level:	<u>low</u>	sdg no.:	<u>Q1782</u>
contract:	<u>LOCK01</u>	lab code:	<u>CHEM</u>	case no.:	<u>Q1782</u>
matrix:	<u>Water</u>	sample id:	<u>Q1782-07</u>	client id:	<u>MW-4MSD</u>
Percent Solids for Sample:	<u>NA</u>	Spiked ID:	<u>Q1782-07MSD</u>	Percent Solids for Spike Sample:	<u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Antimony	ug/L	75 - 125	377		25.0	U	400	94		P
Arsenic	ug/L	75 - 125	350		10.0	U	400	88		P
Barium	ug/L	75 - 125	195		113		100	82		P
Beryllium	ug/L	75 - 125	80.8		3.00	U	100	81		P
Cadmium	ug/L	75 - 125	84.8		3.00	U	100	85		P
Chromium	ug/L	75 - 125	166		5.00	U	200	83		P
Cobalt	ug/L	75 - 125	82.0		15.0	U	100	82		P
Copper	ug/L	75 - 125	120		10.0	U	150	80		P
Iron	ug/L	75 - 125	1460		67.2		1500	93		P
Lead	ug/L	75 - 125	404		6.00	U	500	81		P
Manganese	ug/L	75 - 125	112		98.1		100	14	N	P
Nickel	ug/L	75 - 125	204		20.0	U	250	82		P
Potassium	ug/L	75 - 125	16600		12200		5000	89		P
Selenium	ug/L	75 - 125	836		10.0	U	1000	84		P
Silver	ug/L	75 - 125	31.9		5.00	U	37.5	85		P
Sodium	ug/L	75 - 125	201000		202000		1500	-25		P
Thallium	ug/L	75 - 125	776		20.0	U	1000	78		P
Vanadium	ug/L	75 - 125	121		20.0	U	150	81		P
Zinc	ug/L	75 - 125	96.0		9.56	J	100	86		P

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q1782
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q1782 **SAS No.:** Q1782
Matrix: Water **Level:** LOW **Client ID:** MW-4A
Sample ID: Q1782-07 **Spiked ID:** Q1782-07A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Manganese	ug/L	75 - 125	107		98.1		100	9		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc. **Level:** LOW **SDG No.:** Q1782
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q1782 **SAS No.:** Q1782
Matrix: Water **Sample ID:** Q1782-07 **Client ID:** MW-4DUP
Percent Solids for Sample: NA **Duplicate ID** Q1782-07DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	ug/L	20	25.0	U	25.0	U			P
Arsenic	ug/L	20	10.0	U	10.0	U			P
Barium	ug/L	20	113		114		1		P
Beryllium	ug/L	20	3.00	U	3.00	U			P
Cadmium	ug/L	20	3.00	U	3.00	U			P
Chromium	ug/L	20	5.00	U	5.00	U			P
Cobalt	ug/L	20	15.0	U	15.0	U			P
Copper	ug/L	20	10.0	U	10.0	U			P
Iron	ug/L	20	67.2		66.4		1		P
Lead	ug/L	20	6.00	U	6.00	U			P
Manganese	ug/L	20	98.1		28.2		111	*	P
Nickel	ug/L	20	20.0	U	20.0	U			P
Potassium	ug/L	20	12200		12400		2		P
Selenium	ug/L	20	10.0	U	10.0	U			P
Silver	ug/L	20	5.00	U	5.00	U			P
Sodium	ug/L	20	202000		204000		1		P
Thallium	ug/L	20	20.0	U	20.0	U			P
Vanadium	ug/L	20	20.0	U	20.0	U			P
Zinc	ug/L	20	9.56	J	9.63	J	1		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc. **Level:** LOW **SDG No.:** Q1782
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q1782 **SAS No.:** Q1782
Matrix: Water **Sample ID:** Q1782-07MS **Client ID:** MW-4MSD
Percent Solids for Sample: NA **Duplicate ID** Q1782-07MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	ug/L	20	350		377		7		P
Arsenic	ug/L	20	347		350		1		P
Barium	ug/L	20	191		195		2		P
Beryllium	ug/L	20	77.8		80.8		4		P
Cadmium	ug/L	20	84.7		84.8		0		P
Chromium	ug/L	20	163		166		2		P
Cobalt	ug/L	20	81.6		82.0		0		P
Copper	ug/L	20	119		120		1		P
Iron	ug/L	20	1330		1460		9		P
Lead	ug/L	20	404		404		0		P
Manganese	ug/L	20	106		112		6		P
Nickel	ug/L	20	204		204		0		P
Potassium	ug/L	20	16400		16600		1		P
Selenium	ug/L	20	832		836		0		P
Silver	ug/L	20	31.2		31.9		2		P
Sodium	ug/L	20	198000		201000		2		P
Thallium	ug/L	20	780		776		1		P
Vanadium	ug/L	20	119		121		2		P
Zinc	ug/L	20	93.8		96.0		2		P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q1782
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q1782 **SAS No.:** Q1782

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167583BS							
Antimony	ug/L	400	388		97	80 - 120	P
Arsenic	ug/L	400	375		94	80 - 120	P
Barium	ug/L	100	87.9		88	80 - 120	P
Beryllium	ug/L	100	91.5		92	80 - 120	P
Cadmium	ug/L	100	90.7		91	80 - 120	P
Chromium	ug/L	200	183		92	80 - 120	P
Cobalt	ug/L	100	87.4		87	80 - 120	P
Copper	ug/L	150	140		93	80 - 120	P
Iron	ug/L	1500	1360		91	80 - 120	P
Lead	ug/L	500	453		91	80 - 120	P
Manganese	ug/L	100	90.0		90	80 - 120	P
Nickel	ug/L	250	221		88	80 - 120	P
Potassium	ug/L	5000	4370		87	80 - 120	P
Selenium	ug/L	1000	966		97	80 - 120	P
Silver	ug/L	37.5	34.3		92	80 - 120	P
Sodium	ug/L	1500	1270		85	80 - 120	P
Thallium	ug/L	1000	953		95	80 - 120	P
Vanadium	ug/L	150	132		88	80 - 120	P
Zinc	ug/L	100	94.0		94	80 - 120	P

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

MW-4L

Lab Name: Chemtech Consulting Group

Contract: LOCK01

Lab Code: CHEM Lb No.: lb135445 Lab Sample ID : Q1782-07L SDG No.: Q1782

Matrix (soil/water): Water Level (low/med): LOW

Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Differ-ence	Q	M
		C		C			
Antimony	25.0	U	125	U			P
Arsenic	10.0	U	50.0	U			P
Barium	113		111	J	3		P
Beryllium	3.00	U	15.0	U			P
Cadmium	3.00	U	15.0	U			P
Chromium	5.00	U	25.0	U			P
Cobalt	15.0	U	75.0	U			P
Copper	10.0	U	50.0	U			P
Iron	67.2		250	U	100.0		P
Lead	6.00	U	30.0	U			P
Manganese	98.1		94.7		3		P
Nickel	20.0	U	100	U			P
Potassium	12200		11400		6		P
Selenium	10.0	U	50.0	U			P
Silver	5.00	U	25.0	U			P
Sodium	202000		196000		3		P
Thallium	20.0	U	100	U			P
Vanadium	20.0	U	100	U			P
Zinc	9.56	J	9.46	J	1		P



METAL PREPARATION & INSTRUMENT DATA

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q1782

Contract: LOCK01

Lab Code: CHEM

Case No.: Q1782

SAS No.: Q1782

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Al	Ca	Fe	Mg	Ag
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	-0.0000440	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000930	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0007850	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	-0.0000920	0.0000000	0.0000380	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	-0.0001440	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	-0.0001490	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0001050	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q1782

Contract: LOCK01

Lab Code: CHEM

Case No.: Q1782

SAS No.: Q1782

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		As	Ba	Be	Cd	Co
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0002870
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0000000	0.0009530
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	-0.0039600
Lead	220.353	0.0000000	0.0003170	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	-0.0003570
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0054900
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q1782

Contract: LOCK01

Lab Code: CHEM

Case No.: Q1782

SAS No.: Q1782

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Cr	Cu	K	Mn	Mo
Antimony	206.833	0.0122000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	-0.0029000	0.0000000	0.0000000	0.0000000	0.0004900
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	-0.0000710	-0.0003400
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000070	0.0002200	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	-0.0007860
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0006510	0.0020500
Iron	240.488	0.0000000	0.0000000	0.0000730	0.0000000	-0.0015250
Lead	220.353	0.0000000	0.0000000	0.0000000	0.0001400	-0.0008600
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0007460	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	-0.0000120
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0017400	-0.0100400
Vanadium	292.402	-0.0025100	0.0000000	0.0000000	0.0000000	-0.0072000
Zinc	213.800	0.0000000	0.0009010	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q1782

Contract: LOCK01

Lab Code: CHEM

Case No.: Q1782

SAS No.: Q1782

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Na	Ni	Pb	Sb	Se
Antimony	206.833	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cobalt	228.616	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	-0.0047000	0.0036100	0.0000000	0.0000000
Iron	240.488	0.0000000	-0.0017000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	0.0006580	0.0000000	0.0000000	0.0001290
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0003330	0.0000000	0.0000000
Silver	328.068	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Vanadium	292.402	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0067600	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: Q1782

Contract: LOCK01

Lab Code: CHEM

Case No.: Q1782

SAS No.: Q1782

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Sn	Ti	Tl	V	Zn
Antimony	206.833	-0.0035600	-0.0007970	0.0000000	-0.0018900	0.0000000
Arsenic	193.759	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Barium	493.409	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Beryllium	234.861	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Cadmium	226.502	0.0000000	0.0000630	0.0001280	0.0000000	0.0000000
Chromium	267.716	0.0000000	0.0000000	0.0000000	0.0001110	0.0000000
Cobalt	228.616	0.0000000	0.0018800	0.0000000	0.0000000	0.0000000
Copper	224.700	0.0000000	0.0003840	0.0000000	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	-0.0003610	0.0000000	0.0000000	0.0000000
Manganese	257.610	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Nickel	231.604	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Potassium	766.490	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Selenium	196.090	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Silver	328.068	0.0000000	-0.0007420	0.0000000	0.0000000	0.0000000
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Thallium	190.856	0.0000000	-0.0039700	0.0000000	-0.0115600	0.0000000
Vanadium	292.402	0.0000000	0.0005320	0.0000000	0.0000000	0.0000000
Zinc	213.800	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

METAL PREPARATION & ANALYICAL SUMMARY

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client:	<u>Lockwood, Kessler & Bartlett, Inc.</u>	SDG No.:	<u>Q1782</u>
Contract:	<u>LOCK01</u>	Lab Code:	<u>CHEM</u>
		Method:	<u></u>
		Case No.:	<u>Q1782</u>
		SAS No.:	<u>Q1782</u>

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB167583							
PB167583BL	PB167583BL	MB	WATER	04/14/2025	50.0	25.0	
PB167583BS	PB167583BS	LCS	WATER	04/14/2025	50.0	25.0	
Q1782-01	MW-1	SAM	WATER	04/14/2025	50.0	25.0	
Q1782-02	MW-1	SAM	WATER	04/14/2025	50.0	25.0	
Q1782-03	MW-2	SAM	WATER	04/14/2025	50.0	25.0	
Q1782-04	MW-2	SAM	WATER	04/14/2025	50.0	25.0	
Q1782-05	MW-3	SAM	WATER	04/14/2025	50.0	25.0	
Q1782-06	MW-3	SAM	WATER	04/14/2025	50.0	25.0	
Q1782-07	MW-4	SAM	WATER	04/14/2025	50.0	25.0	
Q1782-07DUP	MW-4DUP	DUP	WATER	04/14/2025	50.0	25.0	
Q1782-07MS	MW-4MS	MS	WATER	04/14/2025	50.0	25.0	
Q1782-07MSD	MW-4MSD	MSD	WATER	04/14/2025	50.0	25.0	
Q1782-08	MW-4	SAM	WATER	04/14/2025	50.0	25.0	

metals
- 14 -
ANALYSIS RUN LOG

Client: Lockwood, Kessler & Bartlett, Inc.

Contract: LOCK01

Lab code: CHEM

Case no.: Q1782

Sas no.: Q1782

Sdg no.: Q1782

Instrument id number:

Method:

Run number: lb135445

Start date: 04/15/2025

End date: 04/15/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1508	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S1	S1	1	1512	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S2	S2	1	1516	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S3	S3	1	1521	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S4	S4	1	1525	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
S5	S5	1	1529	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICV01	ICV01	1	1547	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
LLICV	LLICV	1	1551	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICB01	ICB01	1	1555	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CRI1	CRI1	1	1605	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSA01	ICSA01	1	1609	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
ICSAB01	ICSAB01	1	1613	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV01	CCV01	1	1626	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB01	CCB01	1	1630	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV02	CCV02	1	1716	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB02	CCB02	1	1720	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV03	CCV03	1	1807	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB03	CCB03	1	1811	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCV04	CCV04	1	1908	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB04	CCB04	1	1912	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
PB167583BL	PB167583BL	1	1933	Ag,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
PB167583BS	PB167583BS	1	1938	Ag,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q1782-01	MW-1	1	1950	Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q1782-02	MW-1	1	1954	Fe,Mn
CCV05	CCV05	1	1959	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB05	CCB05	1	2003	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q1782-03	MW-2	1	2007	Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q1782-04	MW-2	1	2011	Fe,Mn
Q1782-05	MW-3	1	2016	Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q1782-06	MW-3	1	2021	Fe,Mn
Q1782-07	MW-4	1	2025	Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q1782-07DUP	MW-4DUP	1	2029	Ag,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q1782-07L	MW-4L	5	2034	Ag,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q1782-07MS	MW-4MS	1	2038	Ag,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q1782-07MSD	MW-4MSD	1	2042	Ag,As,Ba,Be,Cd,Co,Cr,Cu,Fe,K,Mn,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q1782-07A	MW-4A	1	2046	Mn
CCV06	CCV06	1	2051	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB06	CCB06	1	2055	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
Q1782-08	MW-4	1	2059	Fe,Mn
CCV07	CCV07	1	2143	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB07	CCB07	1	2147	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn

metals
- 14 -
ANALYSIS RUN LOG

Client: Lockwood, Kessler & Bartlett, Inc.

Contract: LOCK01

Lab code: CHEM **Case no.:** Q1782 **Sas no.:** Q1782

Sdg no.: Q1782

Instrument id number: **Method:**

Run number: lb135445

Start date: 04/15/2025 **End date:** 04/15/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
CCV08	CCV08	1	2225	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn
CCB08	CCB08	1	2229	Fe,Mn,Ag,As,Ba,Be,Cd,Co,Cr,Cu,K,Na,Ni,Pb,Sb,Se,Tl,V,Zn