



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q1782
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q1782 **SAS No.:** Q1782
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI1	Antimony	54.2	50.0	108	40 - 160	P	04/15/2025	16:05	lb135445
	Arsenic	23.2	20.0	116	40 - 160	P	04/15/2025	16:05	lb135445
	Barium	93.9	100	94	40 - 160	P	04/15/2025	16:05	lb135445
	Beryllium	5.90	6.0	98	40 - 160	P	04/15/2025	16:05	lb135445
	Cadmium	5.84	6.0	97	40 - 160	P	04/15/2025	16:05	lb135445
	Chromium	9.65	10.0	97	40 - 160	P	04/15/2025	16:05	lb135445
	Cobalt	28.6	30.0	95	40 - 160	P	04/15/2025	16:05	lb135445
	Copper	21.4	20.0	107	40 - 160	P	04/15/2025	16:05	lb135445
	Iron	90.6	100	91	40 - 160	P	04/15/2025	16:05	lb135445
	Lead	12.5	12.0	104	40 - 160	P	04/15/2025	16:05	lb135445
	Manganese	20.9	20.0	104	40 - 160	P	04/15/2025	16:05	lb135445
	Nickel	38.5	40.0	96	40 - 160	P	04/15/2025	16:05	lb135445
	Potassium	2000	2000	100	40 - 160	P	04/15/2025	16:05	lb135445
	Selenium	22.6	20.0	113	40 - 160	P	04/15/2025	16:05	lb135445
	Silver	10.3	10.0	103	40 - 160	P	04/15/2025	16:05	lb135445
	Sodium	1770	2000	88	40 - 160	P	04/15/2025	16:05	lb135445
	Thallium	43.2	40.0	108	40 - 160	P	04/15/2025	16:05	lb135445
	Vanadium	39.3	40.0	98	40 - 160	P	04/15/2025	16:05	lb135445
	Zinc	41.4	40.0	104	40 - 160	P	04/15/2025	16:05	lb135445