

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc. **Level:** LOW **SDG No.:** Q1782
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q1782 **SAS No.:** Q1782
Matrix: Water **Sample ID:** Q1782-07 **Client ID:** MW-4DUP
Percent Solids for Sample: NA **Duplicate ID** Q1782-07DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	ug/L	20	25.0	U	25.0	U			P
Arsenic	ug/L	20	10.0	U	10.0	U			P
Barium	ug/L	20	113		114		1		P
Beryllium	ug/L	20	3.00	U	3.00	U			P
Cadmium	ug/L	20	3.00	U	3.00	U			P
Chromium	ug/L	20	5.00	U	5.00	U			P
Cobalt	ug/L	20	15.0	U	15.0	U			P
Copper	ug/L	20	10.0	U	10.0	U			P
Iron	ug/L	20	67.2		66.4		1		P
Lead	ug/L	20	6.00	U	6.00	U			P
Manganese	ug/L	20	98.1		28.2		111	*	P
Nickel	ug/L	20	20.0	U	20.0	U			P
Potassium	ug/L	20	12200		12400		2		P
Selenium	ug/L	20	10.0	U	10.0	U			P
Silver	ug/L	20	5.00	U	5.00	U			P
Sodium	ug/L	20	202000		204000		1		P
Thallium	ug/L	20	20.0	U	20.0	U			P
Vanadium	ug/L	20	20.0	U	20.0	U			P
Zinc	ug/L	20	9.56	J	9.63	J	1		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc. **Level:** LOW **SDG No.:** Q1782
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q1782 **SAS No.:** Q1782
Matrix: Water **Sample ID:** Q1782-07MS **Client ID:** MW-4MSD
Percent Solids for Sample: NA **Duplicate ID** Q1782-07MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	ug/L	20	350		377		7		P
Arsenic	ug/L	20	347		350		1		P
Barium	ug/L	20	191		195		2		P
Beryllium	ug/L	20	77.8		80.8		4		P
Cadmium	ug/L	20	84.7		84.8		0		P
Chromium	ug/L	20	163		166		2		P
Cobalt	ug/L	20	81.6		82.0		0		P
Copper	ug/L	20	119		120		1		P
Iron	ug/L	20	1330		1460		9		P
Lead	ug/L	20	404		404		0		P
Manganese	ug/L	20	106		112		6		P
Nickel	ug/L	20	204		204		0		P
Potassium	ug/L	20	16400		16600		1		P
Selenium	ug/L	20	832		836		0		P
Silver	ug/L	20	31.2		31.9		2		P
Sodium	ug/L	20	198000		201000		2		P
Thallium	ug/L	20	780		776		1		P
Vanadium	ug/L	20	119		121		2		P
Zinc	ug/L	20	93.8		96.0		2		P