

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Lockwood, Kessler & Bartlett, Inc. **SDG No.:** Q1782
Contract: LOCK01 **Lab Code:** CHEM **Case No.:** Q1782 **SAS No.:** Q1782

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167583BS							
Antimony	ug/L	400	388		97	80 - 120	P
Arsenic	ug/L	400	375		94	80 - 120	P
Barium	ug/L	100	87.9		88	80 - 120	P
Beryllium	ug/L	100	91.5		92	80 - 120	P
Cadmium	ug/L	100	90.7		91	80 - 120	P
Chromium	ug/L	200	183		92	80 - 120	P
Cobalt	ug/L	100	87.4		87	80 - 120	P
Copper	ug/L	150	140		93	80 - 120	P
Iron	ug/L	1500	1360		91	80 - 120	P
Lead	ug/L	500	453		91	80 - 120	P
Manganese	ug/L	100	90.0		90	80 - 120	P
Nickel	ug/L	250	221		88	80 - 120	P
Potassium	ug/L	5000	4370		87	80 - 120	P
Selenium	ug/L	1000	966		97	80 - 120	P
Silver	ug/L	37.5	34.3		92	80 - 120	P
Sodium	ug/L	1500	1270		85	80 - 120	P
Thallium	ug/L	1000	953		95	80 - 120	P
Vanadium	ug/L	150	132		88	80 - 120	P
Zinc	ug/L	100	94.0		94	80 - 120	P