

Surrogate Summary

SDG No.: Q1782

Client: Lockwood, Kessler & Bartlett, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1782-01	MW-1	1,2-Dichloroethane-d4	50	53.3	107	74	125
		Dibromofluoromethane	50	52.0	104	75	124
		Toluene-d8	50	50.9	102	86	113
		4-Bromofluorobenzene	50	50.6	101	77	121
Q1782-03	MW-2	1,2-Dichloroethane-d4	50	55.3	111	74	125
		Dibromofluoromethane	50	51.2	102	75	124
		Toluene-d8	50	51.5	103	86	113
		4-Bromofluorobenzene	50	51.5	103	77	121
Q1782-05	MW-3	1,2-Dichloroethane-d4	50	55.1	110	74	125
		Dibromofluoromethane	50	52.0	104	75	124
		Toluene-d8	50	51.0	102	86	113
		4-Bromofluorobenzene	50	52.0	104	77	121
Q1782-09	TRIP BLANK	1,2-Dichloroethane-d4	50	55.2	110	74	125
		Dibromofluoromethane	50	51.6	103	75	124
		Toluene-d8	50	50.8	102	86	113
		4-Bromofluorobenzene	50	52.3	105	77	121
VX0411WBL01	VX0411WBL01	1,2-Dichloroethane-d4	50	53.5	107	74	125
		Dibromofluoromethane	50	51.2	102	75	124
		Toluene-d8	50	50.0	100	86	113
		4-Bromofluorobenzene	50	50.4	101	77	121
VX0411WBS01	VX0411WBS01	1,2-Dichloroethane-d4	50	50.5	101	74	125
		Dibromofluoromethane	50	50.6	101	75	124
		Toluene-d8	50	49.2	98	86	113
		4-Bromofluorobenzene	50	50.6	101	77	121
VX0411WBSD0	VX0411WBSD01	1,2-Dichloroethane-d4	50	52.3	105	74	125
		Dibromofluoromethane	50	49.8	100	75	124
		Toluene-d8	50	49.3	99	86	113
		4-Bromofluorobenzene	50	52.9	106	77	121