

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1782

Client: Lockwood, Kessler & Bartlett, Inc.

Analytical Method: SW8260-Low

Datafile : VX045734.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0411WBSD01	Vinyl chloride	20	17.6	ug/L	88	0		65	117	20
	Chloroethane	20	19.0	ug/L	95	0		56	128	20
	Trichlorofluoromethane	20	19.3	ug/L	97	5		73	115	20
	1,1-Dichloroethene	20	17.9	ug/L	90	1		74	110	20
	Acrylonitrile	100	99.0	ug/L	99	6		73	113	20
	Acetone	100	100	ug/L	100	8		60	125	20
	Carbon disulfide	20	14.5	ug/L	73	3		64	112	20
	Methylene Chloride	20	19.4	ug/L	97	6		72	114	20
	Vinyl Acetate	100	98.0	ug/L	98	6		76	115	20
	2-Butanone	100	100	ug/L	100	4		65	122	20
	Carbon Tetrachloride	20	18.9	ug/L	95	2		77	113	20
	cis-1,2-Dichloroethene	20	18.4	ug/L	92	2		77	110	20
	Bromochloromethane	20	25.7	ug/L	129	5	*	70	124	20
	Chloroform	20	19.7	ug/L	99	7		79	113	20
	1,1,1-Trichloroethane	20	19.3	ug/L	97	4		80	108	20
	Benzene	20	18.7	ug/L	94	3		82	109	20
	1,2-Dichloropropane	20	19.2	ug/L	96	2		83	111	20
	Dibromomethane	20	19.7	ug/L	99	6		82	110	20
	Bromodichloromethane	20	19.5	ug/L	98	6		83	110	20
	4-Methyl-2-Pentanone	100	110	ug/L	110	12		74	118	20
	Toluene	20	19.1	ug/L	96	5		82	110	20
	t-1,3-Dichloropropene	20	18.4	ug/L	92	4		79	110	20
	cis-1,3-Dichloropropene	20	19.9	ug/L	100	5		82	110	20
	2-Hexanone	100	110	ug/L	110	10		73	117	20
	Dibromochloromethane	20	19.2	ug/L	96	3		82	110	20
	1,2-Dibromoethane	20	20.1	ug/L	101	8		81	110	20
	Tetrachloroethene	20	19.2	ug/L	96	0		67	123	20
	Chlorobenzene	20	19.6	ug/L	98	3		82	109	20
	1,1,1,2-Tetrachloroethane	20	19.2	ug/L	96	3		84	111	20
	Ethyl Benzene	20	19.5	ug/L	98	3		83	109	20
	m/p-Xylenes	40	38.9	ug/L	97	3		82	110	20
	o-Xylene	20	19.7	ug/L	99	3		83	109	20
	Styrene	20	20.1	ug/L	101	4		80	111	20
	Bromoform	20	18.8	ug/L	94	4		79	109	20
	1,1,2,2-Tetrachloroethane	20	19.1	ug/L	96	1		76	118	20
	1,2,3-Trichloropropane	20	19.4	ug/L	97	0		75	112	20
	1,4-Dichlorobenzene	20	18.6	ug/L	93	2		82	107	20
	1,2-Dichlorobenzene	20	19.4	ug/L	97	0		82	109	20
	1,2-Dibromo-3-Chloropropane	20	20.3	ug/L	102	3		68	112	20
	Methyl iodide	20	17.6	ug/L	88	8		70	130	20
	trans-1,4-Dichloro-2-butene	20	17.0	ug/L	85	1		79	102	20