

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: LOCK01
 Lab Code: CHEM Case No.: Q1782 SAS No.: Q1782 SDG NO.: Q1782
 Lab File ID: VX045730.D Date Analyzed: 04/11/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:41
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	97283	5.54	164714	6.76	147741	10.05
UPPER LIMIT	194566	6.044	329428	7.257	295482	10.549
LOWER LIMIT	48641.5	5.044	82357	6.257	73870.5	9.549
EPA SAMPLE NO.						
MW-1	68722	5.54	132692	6.76	121469	10.05
MW-2	65299	5.54	128436	6.76	119025	10.06
MW-3	63602	5.55	124448	6.76	115160	10.06
TRIP BLANK	62495	5.55	122581	6.76	112879	10.05
VX0411WBL01	65554	5.55	128675	6.76	118153	10.05
VX0411WBS01	92540	5.54	163903	6.76	140888	10.05
VX0411WBSD01	84320	5.54	150272	6.76	131860	10.05

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.