

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: LOCK01
 Lab Code: CHEM Case No.: Q1782 SAS No.: Q1782 SDG NO.: Q1782
 Lab File ID: VX045730.D Date Analyzed: 04/11/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:41
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	71449	12.018				
UPPER LIMIT	142898	12.518				
LOWER LIMIT	35724.5	11.518				
EPA SAMPLE NO.						
MW-1	48933	12.02				
MW-2	49214	12.02				
MW-3	48473	12.02				
TRIP BLANK	48954	12.02				
VX0411WBL01	48353	12.02				
VX0411WBS01	64715	12.02				
VX0411WBSD01	63206	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.