



## **Report of Analysis**

Client: PSEG Date Collected: 04/11/25

Project: Date Received: 04/11/25

Client Sample ID: 41 MORRIS AVE-NEWARK SDG No.: Q1786

Lab Sample ID: Q1786-01 Matrix: SOIL

Analytical Method: SW8082A % Solid: 100 Decanted:

Sample Wt/Vol: 1 Units: g Final Vol: 10000 uL

Soil Aliquot Vol: uL Test: PCB

Extraction Type: Injection Volume:

GPC Factor: 1.0 PH:

Prep Method: SW3541B

 File ID/Qc Batch:
 Dilution:
 Prep Date
 Date Analyzed
 Prep Batch ID

 PP071272.D
 1
 04/15/25 08:30
 04/15/25 12:13
 PB167594

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	0.51	U	0.12	0.51	mg/Kg
11104-28-2	Aroclor-1221	0.51	U	0.12	0.51	mg/Kg
11141-16-5	Aroclor-1232	0.51	U	0.11	0.51	mg/Kg
53469-21-9	Aroclor-1242	290	E	0.12	0.51	mg/Kg
12672-29-6	Aroclor-1248	0.51	U	0.18	0.51	mg/Kg
11097-69-1	Aroclor-1254	0.51	U	0.096	0.51	mg/Kg
37324-23-5	Aroclor-1262	0.51	U	0.15	0.51	mg/Kg
11100-14-4	Aroclor-1268	0.51	U	0.11	0.51	mg/Kg
11096-82-5	Aroclor-1260	0.51	U	0.097	0.51	mg/Kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	15.8		30 - 150	79%	SPK: 20
2051-24-3	Decachlorobiphenyl	15.3		30 - 150	76%	SPK: 20

## Comments:

U = Not Detected

LOO = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit