

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JAC005

Lab Code: CHEM Case No.: Q1790 SAS No.: Q1790 SDG No.: Q1790

Instrument ID: MSVOA_X Calibration Date/Time: 04/14/2025 09:50

Lab File ID: VX045755.D Init. Calib. Date(s): 04/01/2025 04/01/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 17:06 19:02

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Trichloroethene	0.348	0.349		0.29	20
1,2-Dichloroethane-d4	0.914	0.872		-4.59	20
Dibromofluoromethane	0.355	0.354		-0.28	20
Toluene-d8	1.238	1.196		-3.39	20
4-Bromofluorobenzene	0.451	0.461		2.22	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.