

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JAC005

Lab Code: CHEM Case No.: Q1790 SAS No.: Q1790 SDG No.: Q1790

Instrument ID: MSVOA_Y Calibration Date/Time: 04/15/2025 09:03

Lab File ID: VY021882.D Init. Calib. Date(s): 03/27/2025 03/27/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 10:08 12:02

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Trichloroethene	0.377	0.406		7.69	20
1,2-Dichloroethane-d4	0.542	0.514		-5.17	20
Dibromofluoromethane	0.324	0.329		1.54	20
Toluene-d8	1.234	1.225		-0.73	20
4-Bromofluorobenzene	0.417	0.437		4.8	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.