

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX040225

Review By	John Carlone	Review On	4/2/2025 9:42:46 AM
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:39 PM
SubDirectory	VX040225	HP Acquire Method	HP Processing Method 82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133559,VP133563
Initial Calibration Stds	VP133569,VP133570,VP133571,VP133572,VP133573,VP133574
CCC	VP133564,VP133565
Internal Standard/PEM	
ICV/I.BLK	VP133575
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045524.D	01 Apr 2025 16:15		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX045525.D	01 Apr 2025 17:06	%D failed for Comp. #53 in 01PPB	JC/MD	Ok,M
3	VSTDIC005	VSTDIC005	VX045526.D	01 Apr 2025 17:29		JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX045527.D	01 Apr 2025 17:52		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX045528.D	01 Apr 2025 18:15		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX045529.D	01 Apr 2025 18:38		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX045530.D	01 Apr 2025 19:02		JC/MD	Ok,M
8	IBLK	IBLK	VX045531.D	01 Apr 2025 19:25		JC/MD	Ok
9	VSTDICV050	ICVVX040225	VX045532.D	01 Apr 2025 19:48		JC/MD	Ok,M
10	BFB	BFB	VX045533.D	02 Apr 2025 09:30		JC/MD	Ok
11	VSTDIC050	VSTDIC050	VX045534.D	02 Apr 2025 10:02	pH#Lot#V12668	JC/MD	Ok,M
12	VX0402MBL01	VX0402MBL01	VX045535.D	02 Apr 2025 10:30		JC/MD	Ok
13	VX0402WBL01	VX0402WBL01	VX045536.D	02 Apr 2025 10:53		JC/MD	Ok
14	VX0402WBS01	VX0402WBS01	VX045537.D	02 Apr 2025 11:16		JC/MD	Ok,M
15	VX0402WBSD01	VX0402WBSD01	VX045538.D	02 Apr 2025 11:44		JC/MD	Ok,M
16	Q1697-01	MW-19B-72-040125	VX045539.D	02 Apr 2025 12:07	vial A pH<2 Need 200X	JC/MD	Dilution
17	Q1697-02	IW-01-55-040125	VX045540.D	02 Apr 2025 12:30	vial A pH<2 Need 10X	JC/MD	Dilution

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX040225

Review By	John Carlone	Review On	4/2/2025 9:42:46 AM
Supervise By	Amit Patel	Supervise On	4/2/2025 2:14:39 PM
SubDirectory	VX040225	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133559,VP133563		
Initial Calibration Stds	VP133569,VP133570,VP133571,VP133572,VP133573,VP133574		
CCC	VP133564,VP133565		
Internal Standard/PEM			
ICV/I.BLK	VP133575		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	IBLK	IBLK	VX045541.D	02 Apr 2025 12:54		JC/MD	Ok
19	Q1697-01DL	MW-19B-72-040125DL	VX045542.D	02 Apr 2025 13:21	vial B pH<2	JC/MD	Ok
20	Q1697-02DL	IW-01-55-040125DL	VX045543.D	02 Apr 2025 13:44	vial B pH<2	JC/MD	Ok
21	Q1697-03	IW-02-55-040125	VX045544.D	02 Apr 2025 14:07	Need lower dilution	JC/MD	Not Ok
22	Q1697-04	IW-02-55-040125-FD	VX045545.D	02 Apr 2025 14:31	Need lower dilution	JC/MD	Not Ok
23	Q1697-05	IW-03-55-040125	VX045546.D	02 Apr 2025 14:54	Need lower dilution	JC/MD	Not Ok
24	Q1697-07	MW-19B-72-040125	VX045547.D	02 Apr 2025 15:18	Not On Login	JC/MD	Not Ok
25	IBLK	IBLK	VX045548.D	02 Apr 2025 15:41		JC/MD	Ok
26	Q1697-03	IW-02-55-040125	VX045549.D	02 Apr 2025 16:04	vial A pH<2	JC/MD	Ok
27	Q1697-04	IW-02-55-040125-FD	VX045550.D	02 Apr 2025 16:27	vial A pH<2	JC/MD	Ok
28	Q1697-05	IW-03-55-040125	VX045551.D	02 Apr 2025 16:51	vial A pH<2	JC/MD	Ok
29	Q1697-06	TB-01-040125	VX045552.D	02 Apr 2025 17:14	vial A pH<2 TB;Hit of comp.#44	JC/MD	ReRun
30	Q1623-01	Storage-Blank-SOIL-RE	VX045553.D	02 Apr 2025 17:37	vial A pH<2	JC/MD	Ok
31	Q1623-02	Storage-Blank-WATER	VX045554.D	02 Apr 2025 18:01	vial A pH<2	JC/MD	Ok
32	Q1623-03	Storage-Blank-WATER	VX045555.D	02 Apr 2025 18:24	vial A pH<2	JC/MD	Ok
33	Q1623-04	Storage-Blank-SAMPLE	VX045556.D	02 Apr 2025 18:47	vial A pH<2	JC/MD	Ok
34	VSTDCCC050	VSTDCCC050EC	VX045557.D	02 Apr 2025 19:10		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX041125

Review By	John Carlone	Review On	4/14/2025 9:35:27 AM
Supervise By	Maresh Dadoda	Supervise On	4/14/2025 3:43:21 PM
SubDirectory	VX041125	HP Acquire Method	HP Processing Method 82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133640
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133641,VP133642

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045729.D	11 Apr 2025 10:14		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX045730.D	11 Apr 2025 10:41	pH#Lot#V12668	JC/MD	Ok,M
3	VX0411MBL01	VX0411MBL01	VX045731.D	11 Apr 2025 11:09		JC/MD	Ok
4	VX0411WBL01	VX0411WBL01	VX045732.D	11 Apr 2025 11:32		JC/MD	Ok
5	VX0411WBS01	VX0411WBS01	VX045733.D	11 Apr 2025 11:55		JC/MD	Ok,M
6	VX0411WBSD01	VX0411WBSD01	VX045734.D	11 Apr 2025 12:22		JC/MD	Ok,M
7	Q1773-02	BP-TT190D2-GW-2025	VX045735.D	11 Apr 2025 12:45	vial A pH<2	JC/MD	Ok
8	Q1773-03	BP-TT190D1-GW-2025	VX045736.D	11 Apr 2025 13:08	vial A pH<2 Need 10X	JC/MD	Dilution
9	Q1773-01	BP-TB-20250407	VX045737.D	11 Apr 2025 13:32	vial A pH<2 TB;E flag of previous sample	JC/MD	ReRun
10	Q1790-17	EB01-041025	VX045738.D	11 Apr 2025 13:55	vial A pH<2 EB	JC/MD	Ok
11	Q1790-19	FB01-041025	VX045739.D	11 Apr 2025 14:18	vial A pH<2 FB	JC/MD	Ok
12	Q1782-09	TRIP BLANK	VX045740.D	11 Apr 2025 14:41	vial A pH<2 TB	JC/MD	Ok
13	Q1782-01	MW-1	VX045741.D	11 Apr 2025 15:04	vial A pH<2	JC/MD	Ok
14	Q1782-03	MW-2	VX045742.D	11 Apr 2025 15:28	vial A pH<2	JC/MD	Ok
15	Q1782-05	MW-3	VX045743.D	11 Apr 2025 15:51	vial A pH<2	JC/MD	Ok
16	Q1767-01	NWB-2152	VX045744.D	11 Apr 2025 16:14	run @10x	JC/MD	Not Ok
17	VX0411MBS01	VX0411MBS01	VX045745.D	11 Apr 2025 16:37		JC/MD	Ok,M

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX041125

Review By	John Carlone	Review On	4/14/2025 9:35:27 AM
Supervise By	Maresh Dadoda	Supervise On	4/14/2025 3:43:21 PM
SubDirectory	VX041125	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133640		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133641,VP133642		

18	Q1773-03DL	BP-TT190D1-GW-2025	VX045746.D	11 Apr 2025 17:00	vial B pH<2	JC/MD	Ok
19	Q1773-01	BP-TB-20250407	VX045747.D	11 Apr 2025 17:23	vial B pH<2 TB	JC/MD	Ok
20	Q1790-01	S-872-G1-SO-20.0-20.5	VX045748.D	11 Apr 2025 17:47	run st. MEOH	JC/MD	Not Ok
21	Q1790-04	S-871-G1-SO-23.0-23.5	VX045749.D	11 Apr 2025 18:10	run st. MEOH	JC/MD	Not Ok
22	Q1790-06	S-870-G1-SO-22.0-22.5	VX045750.D	11 Apr 2025 18:33	run st. MEOH	JC/MD	Not Ok
23	IBLK	IBLK	VX045751.D	11 Apr 2025 18:56		JC/MD	Ok
24	IBLK	IBLK	VX045752.D	11 Apr 2025 19:19		JC/MD	Ok
25	VSTDCCC050	VSTDCCC050EC	VX045753.D	11 Apr 2025 19:42		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX041425

Review By	Amit Patel	Review On	4/15/2025 2:25:46 PM
Supervise By	Semsettin Yesilyurt	Supervise On	4/15/2025 2:48:46 PM
SubDirectory	VX041425	HP Acquire Method	HP Processing Method 82X040225W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133654
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133655,VP133656

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045754.D	14 Apr 2025 09:20		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX045755.D	14 Apr 2025 09:50	pH#Lot#V12668	JC/MD	Ok,M
3	VX0414MBL01	VX0414MBL01	VX045756.D	14 Apr 2025 10:18		JC/MD	Ok
4	VX0414WBL01	VX0414WBL01	VX045757.D	14 Apr 2025 10:41		JC/MD	Ok
5	VX0414MBS01	VX0414MBS01	VX045758.D	14 Apr 2025 11:04		JC/MD	Ok,M
6	VX0414MBSD01	VX0414MBSD01	VX045759.D	14 Apr 2025 11:31		JC/MD	Ok,M
7	Q1790-01	S-872-G1-SO-20.0-20.5	VX045760.D	14 Apr 2025 11:55	Run soil	JC/MD	Not Ok
8	Q1790-04	S-871-G1-SO-23.0-23.5	VX045761.D	14 Apr 2025 12:18		JC/MD	Ok
9	Q1790-06	S-870-G1-SO-22.0-22.5	VX045762.D	14 Apr 2025 12:41		JC/MD	Ok
10	VX0414WBS01	VX0414WBS01	VX045763.D	14 Apr 2025 13:04		JC/MD	Ok,M
11	VX0414WBSD01	VX0414WBSD01	VX045764.D	14 Apr 2025 13:39		JC/MD	Ok,M
12	Q1767-01	NWB-2152	VX045765.D	14 Apr 2025 14:02	vial A pH#1.0	JC/MD	Ok
13	Q1799-02	FRAC-TANK-216-PPR	VX045766.D	14 Apr 2025 14:25	vial A pH#1.0	JC/MD	Ok
14	PB167573TB	PB167573TB	VX045767.D	14 Apr 2025 14:48		JC/MD	Ok
15	Q1771-04	TP-6C	VX045768.D	14 Apr 2025 15:12	vial A pH#5.0	JC/MD	Ok
16	Q1771-08	TP-7C	VX045769.D	14 Apr 2025 15:35	vial A pH#5.0	JC/MD	Ok
17	Q1771-12	TP-3A	VX045770.D	14 Apr 2025 15:58	vial A pH#5.0	JC/MD	Ok
18	Q1779-04	TP-22	VX045771.D	14 Apr 2025 16:21	vial A pH#5.0	JC/MD	Ok,M

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX041425

Review By	Amit Patel	Review On	4/15/2025 2:25:46 PM
Supervise By	Semsettin Yesilyurt	Supervise On	4/15/2025 2:48:46 PM
SubDirectory	VX041425	HP Acquire Method	HP Processing Method 82X040225W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133654		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133655,VP133656		

19	Q1779-08	TP-23	VX045772.D	14 Apr 2025 16:44	vial A pH#5.0	JC/MD	Ok
20	Q1781-04	WC-13	VX045773.D	14 Apr 2025 17:08	vial A pH#5.0	JC/MD	Ok
21	Q1781-08	WC-12	VX045774.D	14 Apr 2025 17:31	vial A pH#5.0	JC/MD	Ok
22	Q1781-12	WC-11	VX045775.D	14 Apr 2025 17:54	vial A pH#5.0	JC/MD	Ok
23	Q1787-04	TP-5C	VX045776.D	14 Apr 2025 18:17	Not on Login	JC/MD	Not Ok
24	Q1787-08	TP-2C	VX045777.D	14 Apr 2025 18:40	Not on Login	JC/MD	Not Ok
25	Q1787-12	TP-1	VX045778.D	14 Apr 2025 19:03	vial A pH#5.0	JC/MD	Ok
26	Q1788-04	WC-1	VX045779.D	14 Apr 2025 19:26	vial A pH#5.0	JC/MD	Ok
27	Q1798-02	TOTE(COMP)-1799	VX045780.D	14 Apr 2025 19:50	vial A pH#1.0 Need 250X	JC/MD	Dilution
28	Q1795-02	31125	VX045781.D	14 Apr 2025 20:13	Run @ 10x	JC/MD	Not Ok
29	VSTDCCC050	VSTDCCC050EC	VX045782.D	14 Apr 2025 20:36		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY032725

Review By	Mahesh Dadoda	Review On	3/28/2025 3:29:57 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	3/28/2025 3:30:48 PM		
SubDirectory	VY032725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y032725s.m

STD. NAME	STD REF.#
Tune/Reschk	VP133499
Initial Calibration Stds	VP133500,VP133501,VP133502,VP133503,VP133505,VP133508
CCC	VP133510
Internal Standard/PEM	VP131783
ICV/I.BLK	VP133509
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021655.D	27 Mar 2025 09:23		SY/MD	Ok
2	VSTDIC005	VSTDIC005	VY021656.D	27 Mar 2025 10:08		SY/MD	Ok,M
3	VSTDIC010	VSTDIC010	VY021657.D	27 Mar 2025 10:31		SY/MD	Ok,M
4	VSTDIC020	VSTDIC020	VY021658.D	27 Mar 2025 10:54	Comp.#20 is on Linear Regression	SY/MD	Ok,M
5	VSTDIC050	VSTDIC050	VY021659.D	27 Mar 2025 11:16		SY/MD	Ok,M
6	VSTDIC100	VSTDIC100	VY021660.D	27 Mar 2025 11:39		SY/MD	Ok,M
7	VSTDIC150	VSTDIC150	VY021661.D	27 Mar 2025 12:02		SY/MD	Ok,M
8	VIBLK	VIBLK	VY021662.D	27 Mar 2025 12:35		SY/MD	Ok
9	VSTDICV050	ICVVY032725	VY021663.D	27 Mar 2025 13:15		SY/MD	Ok,M
10	VY0327SBL01	VY0327SBL01	VY021664.D	27 Mar 2025 13:59		SY/MD	Ok
11	VY0327SBS01	VY0327SBS01	VY021665.D	27 Mar 2025 14:38		SY/MD	Ok,M
12	VY0327SBS01	VY0327SBS01	VY021666.D	27 Mar 2025 15:00		SY/MD	Ok,M
13	Q1661-02	351	VY021667.D	27 Mar 2025 15:24	vial-B Not purge	SY/MD	Not Ok
14	Q1645-03RE	TP-4-VOCRE	VY021668.D	27 Mar 2025 15:48	vial-B Internal Standard Fail	SY/MD	Confirms
15	Q1662-03	TP-6-VOC	VY021669.D	27 Mar 2025 16:11	vial-A NOT PURGE	SY/MD	Not Ok
16	Q1662-07	TP-7-VOC	VY021670.D	27 Mar 2025 16:35	vial-A Internal Standard Fail	SY/MD	ReRun
17	Q1661-02	351	VY021671.D	27 Mar 2025 16:58	vial-A NOT PURGE	SY/MD	Not Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY032725

Review By	Mahesh Dadoda	Review On	3/28/2025 3:29:57 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	3/28/2025 3:30:48 PM		
SubDirectory	VY032725	HP Acquire Method	MSVOA_Y	HP Processing Method	82y032725s.m
STD. NAME		STD REF.#			
Tune/Reschk		VP133499			
Initial Calibration Stds		VP133500,VP133501,VP133502,VP133503,VP133505,VP133508			
CCC		VP133510			
Internal Standard/PEM		VP131783			
ICV/I.BLK		VP133509			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q1650-02	STOCK-PILE-BIN2	VY021672.D	27 Mar 2025 17:21	vial-A NOT PURGE	SY/MD	Not Ok
19	Q1654-01	RT3407	VY021673.D	27 Mar 2025 17:45	vial-A Internal Standard Fail	SY/MD	Ok
20	Q1656-01	VNJ239	VY021674.D	27 Mar 2025 18:08	vial-A NOT PURGE	SY/MD	Not Ok
21	Q1648-09	TP-7	VY021675.D	27 Mar 2025 18:32	vial-A	SY/MD	Ok
22	Q1648-07	TP-6	VY021676.D	27 Mar 2025 18:55	vial-A	SY/MD	Ok
23	Q1648-05	TP-3	VY021677.D	27 Mar 2025 19:19	vial-A	SY/MD	Ok
24	Q1648-01	TP-4	VY021678.D	27 Mar 2025 19:42	vial-A	SY/MD	Ok
25	Q1648-03	TP-5	VY021679.D	27 Mar 2025 20:05	vial-A	SY/MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VY021680.D	27 Mar 2025 20:28		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY041525

Review By	Maresh Dadoda	Review On	4/16/2025 3:42:41 PM
Supervise By	Semsettin Yesilyurt	Supervise On	4/16/2025 3:46:30 PM
SubDirectory	VY041525	HP Acquire Method	MSVOA_Y HP Processing Method 82y032725s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133660
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133661,VP133662 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021881.D	15 Apr 2025 08:33		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021882.D	15 Apr 2025 09:03		SY/MD	Ok,M
3	VY0415SBL01	VY0415SBL01	VY021883.D	15 Apr 2025 09:37		SY/MD	Ok
4	VY0415SBS01	VY0415SBS01	VY021884.D	15 Apr 2025 10:08		SY/MD	Ok,M
5	VY0415SBSD01	VY0415SBSD01	VY021885.D	15 Apr 2025 10:31		SY/MD	Ok,M
6	Q1790-01	S-872-G1-SO-20.0-20.5	VY021886.D	15 Apr 2025 11:18	vial-A	SY/MD	Ok
7	Q1792-01RE	TR-05-041125RE	VY021887.D	15 Apr 2025 11:42	vial-B Internal Standard Fail	SY/MD	Confirms
8	Q1785-03	HR-03-041125	VY021888.D	15 Apr 2025 12:05	vial-B Not purge	SY/MD	Not Ok
9	Q1806-01	OR-03-041425	VY021889.D	15 Apr 2025 12:29	vial-A	SY/MD	Ok
10	Q1807-01	OK-02-041425	VY021890.D	15 Apr 2025 12:52	vial-A	SY/MD	Ok
11	Q1808-01	OILY-SOIL-PILE	VY021891.D	15 Apr 2025 13:16	vial-A	SY/MD	Ok
12	Q1808-03	LAW-25-0060	VY021892.D	15 Apr 2025 13:39	vial-A	SY/MD	Ok
13	Q1790-05	S-872-G1-SO-20.0-20.5	VY021893.D	15 Apr 2025 14:02	vial-A	SY/MD	Ok
14	IBLK	TP-15-VOC	VY021894.D	15 Apr 2025 14:26		SY/MD	Ok
15	Q1790-02MS	S-872-G1-SO-20.0-20.5	VY021895.D	15 Apr 2025 14:49	vial-A	SY/MD	Ok,M
16	Q1790-03MSD	S-872-G1-SO-20.0-20.5	VY021896.D	15 Apr 2025 15:11	vial-A Internal Standard Fail	SY/MD	Ok,M
17	IBLK	IBLK	VY021897.D	15 Apr 2025 15:35		SY/MD	Ok
18	VSTDCCC050	VSTDCCC050EC	VY021898.D	15 Apr 2025 15:57		SY/MD	Ok,M

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QCBatch ID # VY041525

Review By	Mahesh Dadoda	Review On	4/16/2025 3:42:41 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	4/16/2025 3:46:30 PM		
SubDirectory	VY041525	HP Acquire Method	MSVOA_Y	HP Processing Method	82y032725s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133660			
CCC		VP133661,VP133662			
Internal Standard/PEM		VP131783			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

M : Manual Integration