

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: Q1790 SAS No.: Q1790 SDG NO.: Q1790
Lab File ID: VX045730.D Date Analyzed: 04/11/2025
Instrument ID: MSVOA_X Time Analyzed: 10:41
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	97283	5.54	164714	6.76	147741	10.05
UPPER LIMIT	194566	6.044	329428	7.257	295482	10.549
LOWER LIMIT	48641.5	5.044	82357	6.257	73870.5	9.549
EPA SAMPLE NO.						
EB01-041025	68658	5.54	132687	6.76	124480	10.05
FB01-041025	62833	5.54	124103	6.76	115712	10.06
VX0411WBL01	65554	5.55	128675	6.76	118153	10.05
VX0411WBS01	92540	5.54	163903	6.76	140888	10.05
VX0411WBSD01	84320	5.54	150272	6.76	131860	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.