

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: Q1790 SAS No.: Q1790 SDG NO.: Q1790
 Lab File ID: VX045755.D Date Analyzed: 04/14/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:50
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	95673	5.54	163598	6.75	144325	10.05
UPPER LIMIT	191346	6.038	327196	7.251	288650	10.549
LOWER LIMIT	47836.5	5.038	81799	6.251	72162.5	9.549
EPA SAMPLE NO.						
S-871-G1-SO-23.0-23.5-041025	72197	5.54	142076	6.76	128869	10.06
S-870-G1-SO-22.0-22.5-041025	67235	5.54	130622	6.76	118702	10.06
VX0414MBL01	66440	5.54	131663	6.76	120808	10.05
VX0414MBS01	93283	5.54	167485	6.76	143860	10.05

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.