

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: Q1790 SAS No.: Q1790 SDG NO.: Q1790
 Lab File ID: VY021882.D Date Analyzed: 04/15/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:03
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	176924	13.346				
UPPER LIMIT	353848	13.846				
LOWER LIMIT	88462	12.846				
EPA SAMPLE NO.						
S-872-G1-SO-20.0-20.5-041025	180946	13.35				
S-872-G1-SO-20.0-20.5-041025M	117205	13.35				
S-872-G1-SO-20.0-20.5-041025M	98584	13.35				
S-871-G1-SO-23.0-23.5-041025	162786	13.35				
VY0415SBL01	163838	13.35				
VY0415SBS01	176232	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.