

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1793 SAS No.: Q1793 SDG NO.: Q1793

EPA Sample No.: SSTD020212 Date Analyzed: 04/15/2025

Lab File ID: BN036913.D Time Analyzed: 12:17

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	80238	7.628	356842	10.42	241823	14.28
UPPER LIMIT	160476	8.128	713684	10.916	483646	14.781
LOWER LIMIT	40119	7.128	178421	9.916	120912	13.781
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	56557	7.63	232029	10.41	155550	14.28

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.