

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1793 SAS No.: Q1793 SDG NO.: Q1793

EPA Sample No.: SSTD020212 Date Analyzed: 04/15/2025

Lab File ID: BN036913.D Time Analyzed: 12:17

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	479431	17.028	450161	21.263	411074	24.151
UPPER LIMIT	958862	17.528	900322	21.763	822148	24.651
LOWER LIMIT	239716	16.528	225081	20.763	205537	23.651
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	319586	17.03	306595	21.26	283789	24.15

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.