

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CHEM02

Lab Code: CHEM

SAS No.: Q1793

SDG NO.: Q1793

Lab File ID: BN036910.D

DFTPP Injection Date: 04/15/2025

Instrument ID: BNA_N

DFTPP Injection Time: 10:20

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	70
68	Less than 2.0% of mass 69	0.5 (0.8) 1
69	Mass 69 relative abundance	58.1
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	52.9
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	25.6
365	Greater than 1% of mass 198	4.2
441	Present, but less than mass 443	8.3
442	Greater than 50% of mass 198	53.9
443	15.0 - 24.0% of mass 442	10.5 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SVOC-GPC2-BLANK	Q1793-10	BN036920.D	04/15/2025	17:30
SSTD020366	SSTDCCC020EC	BN036921.D	04/15/2025	18:09