

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CHEM02

Lab Code: CHEM Case No.: Q1793 SAS No.: Q1793 SDG No.: Q1793

Instrument ID: MSVOA_N Calibration Date/Time: 04/16/2025 09:14

Lab File ID: VN086286.D Init. Calib. Date(s): 04/15/2025 04/15/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:29 14:29

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Acrolein	250.000	262.46	-4.983606	20
Acetone	250.000	253.40	-1.3612	20