

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q1793 SAS No.: Q1793 SDG NO.: Q1793
 Lab File ID: VN086286.D Date Analyzed: 04/16/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:14
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	177295	13.788				
UPPER LIMIT	354590	14.288				
LOWER LIMIT	88647.5	13.288				
EPA SAMPLE NO.						
Storage-Blank-SOIL-REF-6	148157	13.79				
Storage-Blank-WATER-REF-5	126085	13.79				
Storage-Blank-WATER-REF-4	150714	13.79				
Storage-Blank-SAMPLE-MANAGEMENT	136544	13.79				
VN0416WBL01	162383	13.79				
VN0416WBS01	208645	13.79				
VN0416WBSD01	160957	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.