



CASE NARRATIVE

ENTACT

Project Name: 540 Degraw St, Brooklyn, NY - E9309

Project # N/A

Chemtech Project # Q1800 Test Name: TCLP Herbicide

A. Number of Samples and Date of Receipt:

4 Solid samples were received on 04/14/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: ASTM Ammonia, ASTM COD, ASTM Leach Extraction, ASTM Oil and Grease, ASTM TS, Corrosivity, Ignitability, Oil and Grease, Paint Filter, PCB, pH, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL, TS and TVS. This data package contains results for TCLP Herbicide.

C. Analytical Techniques:

The analysis was performed on instrument ECD_S. The front column is RTX-CLPesticides which is 30 meters, 0.32 mm ID, 0. 5 um df,: Catalog # 11139. The rear column is RTX-CLPesticides2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 11324The analysis of TCLP Herbicides was based on method 8151A and extraction was done based on method 3510 and TCLP extraction method was 1311.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for PB167608BL [2,4-DCAA(1) - 136%] these compound did not meet the NJDKQP criteria but met the inhouse criteria

The Retention Times were acceptable for all samples.

The MS {Q1800-03MS} with File ID: PS029830.D recoveries met the requirements for all compounds except for 2,4,5-TP(Silvex)[147%] and 2,4-D[143%] these compound did not meet the NJDKQP criteria and in-house criteria, Due to matrix interference.

The MSD {Q1800-03MSD} with File ID: PS029831.D recoveries met the acceptable requirements except for 2,4,5-TP(Silvex)[147%] and 2,4-D[142%] these compound did not meet the NJDKQP criteria and in-house criteria, Due to matrix interference.





The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

E. Additional Comments:

This data package has been revised due to reporting Unit Changed as per client request.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature	;		