

SOP ID : MSM4500-NH3 B,G-Ammonia-17

SDG No : N/A

Start Digest Date: 04/17/2025 Time : 09:10 Temp : 150 °C

Matrix : WATER

End Digest Date: 04/17/2025 Time : 10:10 Temp : 160 °C

Pipette ID : WC

Balance ID : WC SC-7

Hood ID : HOOD#2

Digestion tube ID : M5595

Block Thermometer ID : WC CYANIDE

Block ID : WC-DIST-BLOCK-1

Filter paper ID : N/A

Prep Technician Signature: RM

Weigh By : RM

pH Meter ID : N/A

Supervisor Signature: 12

Standardized Name	MLS USED	STD REF. # FROM LOG
LCSS	1.0ML	WP112614
MS/MSD SPIKE SOL.	1.0ML	WP112613
PBW	50.0ML	W3112
RL CHECK	0.1ML	WP112613
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
BORATE BUFFER	2.5ML	WP111325
NAOH 6N	0.5-2.0ML	WP111318
H2SO4 0.04N	5.0ML	WP110335
pH Paper 0-14	N/A	W3140
KI-starch paper	N/A	W3155
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

ALL GLASSWEAR ARE STEAMED OUT AND THERE WERE NO TRACE OF AMMONIA USING NESLER REAGENT
WP111604

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
04/17/2025 10:20	RM CWL	RM CWL
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Vol (ml)	Final Vol (ml)	pH	Sulfide	Oxidizing	Nitrate/ Nitrite	Comment	Prep Pos
PB167640BL	PBW640	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
PB167640BS	LCS640	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
PB167640TB	LEB640	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q1800-04DUP	WC-A4-01-CDUP	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q1800-04MS	WC-A4-01-CMS	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q1800-04MSD	WC-A4-01-CMSD	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q1800-04	WC-A4-01-C	50	50	<2	N/A	Negative	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A