



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1800

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135420

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.02	7	100	90-110	04/14/2025
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	04/14/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	04/14/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1800

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135421

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.02	7	100	90-110	04/14/2025
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	04/14/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	04/14/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1800

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135440

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.09	0.099	91	85-115	04/15/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	04/15/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.23	0.25	92	90-110	04/15/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.23	0.25	92	90-110	04/15/2025
Sample ID: CCV4 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	04/15/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1800

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135458

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV ASTM COD	mg/L	49.329	50	99	95-105	01/22/2025
Sample ID: CCV1 ASTM COD	mg/L	49.329	50	99	95-105	04/16/2025
Sample ID: CCV2 ASTM COD	mg/L	48.339	50	97	95-105	04/16/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1800

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135483

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 ASTM Ammonia	mg/L	0.93	1	93	90-110	04/18/2025
Sample ID: CCV1 ASTM Ammonia	mg/L	0.95	1	95	90-110	04/18/2025
Sample ID: CCV2 ASTM Ammonia	mg/L	1	1	100	90-110	04/18/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q1800

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB135483

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------