

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1800

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135450

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135450BS							
Oil and Grease	mg/Kg	100	94.9		95	1	80-120	04/16/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1800

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135454

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135454BS							
ASTM Oil and Grease	mg/L	20.0	16.8		84	1	78-114	04/16/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1800

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135458

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135458BS							
ASTM COD	mg/L	50	50.3		101	1	90-110	04/16/2025

Laboratory Control Sample Summary

Client: ENTACT

SDG No.: Q1800

Project: 540 Degraw St, Brooklyn, NY - E9309

Run No.: LB135483

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167640BS							
ASTM Ammonia	mg/L	1	0.94		94	1	90-110	04/18/2025