



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT SDG No.: Q1800
Contract: ENTA05 Lab Code: CHEM Case No.: Q1800 SAS No.: Q1800
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.20	0.2	98	40 - 160	CV	04/16/2025	13:22	LB135455
CRI	Arsenic	23.3	20.0	116	40 - 160	P	04/16/2025	13:46	LB135463
	Barium	91.3	100	91	40 - 160	P	04/16/2025	13:46	LB135463
	Cadmium	6.00	6.0	100	40 - 160	P	04/16/2025	13:46	LB135463
	Chromium	9.57	10.0	96	40 - 160	P	04/16/2025	13:46	LB135463
	Copper	22.4	20.0	112	40 - 160	P	04/16/2025	13:46	LB135463
	Lead	12.9	12.0	107	40 - 160	P	04/16/2025	13:46	LB135463
	Nickel	39.1	40.0	98	40 - 160	P	04/16/2025	13:46	LB135463
	Selenium	19.0	20.0	95	40 - 160	P	04/16/2025	13:46	LB135463
	Silver	9.95	10.0	100	40 - 160	P	04/16/2025	13:46	LB135463
	Zinc	42.2	40.0	106	40 - 160	P	04/16/2025	13:46	LB135463