



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>ENTACT</u>				SDG No.:	<u>Q1800</u>			
Contract:	<u>ENTA05</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q1800</u>	SAS No.:	<u>Q1800</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB90	Mercury	0.20	+/-0.20	U	0.20	CV	04/16/2025	13:15	LB135455

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q1800

Contract: ENTA05

Lab Code: CHEM

Case No.: Q1800

SAS No.: Q1800

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB79	Mercury	0.20	+/-0.20	U	0.20	CV	04/16/2025	13:20	LB135455
CCB80	Mercury	0.20	+/-0.20	U	0.20	CV	04/16/2025	13:52	LB135455
CCB81	Mercury	0.20	+/-0.20	U	0.20	CV	04/16/2025	14:20	LB135455

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q1800</u>							
Contract: <u>ENTA05</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1800</u>	SAS No.: <u>Q1800</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	04/16/2025	13:42	LB135463
	Barium	100	+/-100	U	100	P	04/16/2025	13:42	LB135463
	Cadmium	6.00	+/-6.00	U	6.00	P	04/16/2025	13:42	LB135463
	Chromium	10.0	+/-10.0	U	10.0	P	04/16/2025	13:42	LB135463
	Copper	20.0	+/-20.0	U	20.0	P	04/16/2025	13:42	LB135463
	Lead	12.0	+/-12.0	U	12.0	P	04/16/2025	13:42	LB135463
	Nickel	40.0	+/-40.0	U	40.0	P	04/16/2025	13:42	LB135463
	Selenium	20.0	+/-20.0	U	20.0	P	04/16/2025	13:42	LB135463
	Silver	10.0	+/-10.0	U	10.0	P	04/16/2025	13:42	LB135463
	Zinc	40.0	+/-40.0	U	40.0	P	04/16/2025	13:42	LB135463

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT		SDG No.: Q1800							
Contract: ENTA05	Lab Code: CHEM	Case No.: Q1800	SAS No.: Q1800						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	04/16/2025	14:12	LB135463
	Barium	100	+/-100	U	100	P	04/16/2025	14:12	LB135463
	Cadmium	6.00	+/-6.00	U	6.00	P	04/16/2025	14:12	LB135463
	Chromium	10.0	+/-10.0	U	10.0	P	04/16/2025	14:12	LB135463
	Copper	20.0	+/-20.0	U	20.0	P	04/16/2025	14:12	LB135463
	Lead	12.0	+/-12.0	U	12.0	P	04/16/2025	14:12	LB135463
	Nickel	40.0	+/-40.0	U	40.0	P	04/16/2025	14:12	LB135463
	Selenium	20.0	+/-20.0	U	20.0	P	04/16/2025	14:12	LB135463
	Silver	10.0	+/-10.0	U	10.0	P	04/16/2025	14:12	LB135463
	Zinc	40.0	+/-40.0	U	40.0	P	04/16/2025	14:12	LB135463
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	04/16/2025	15:23	LB135463
	Barium	100	+/-100	U	100	P	04/16/2025	15:23	LB135463
	Cadmium	6.00	+/-6.00	U	6.00	P	04/16/2025	15:23	LB135463
	Chromium	10.0	+/-10.0	U	10.0	P	04/16/2025	15:23	LB135463
	Copper	20.0	+/-20.0	U	20.0	P	04/16/2025	15:23	LB135463
	Lead	12.0	+/-12.0	U	12.0	P	04/16/2025	15:23	LB135463
	Nickel	40.0	+/-40.0	U	40.0	P	04/16/2025	15:23	LB135463
	Selenium	20.0	+/-20.0	U	20.0	P	04/16/2025	15:23	LB135463
	Silver	10.0	+/-10.0	U	10.0	P	04/16/2025	15:23	LB135463
	Zinc	40.0	+/-40.0	U	40.0	P	04/16/2025	15:23	LB135463
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	04/16/2025	16:14	LB135463
	Barium	100	+/-100	U	100	P	04/16/2025	16:14	LB135463
	Cadmium	6.00	+/-6.00	U	6.00	P	04/16/2025	16:14	LB135463
	Chromium	10.0	+/-10.0	U	10.0	P	04/16/2025	16:14	LB135463
	Copper	20.0	+/-20.0	U	20.0	P	04/16/2025	16:14	LB135463
	Lead	12.0	+/-12.0	U	12.0	P	04/16/2025	16:14	LB135463
	Nickel	40.0	+/-40.0	U	40.0	P	04/16/2025	16:14	LB135463
	Selenium	20.0	+/-20.0	U	20.0	P	04/16/2025	16:14	LB135463
	Silver	10.0	+/-10.0	U	10.0	P	04/16/2025	16:14	LB135463
	Zinc	40.0	+/-40.0	U	40.0	P	04/16/2025	16:14	LB135463
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	04/16/2025	17:06	LB135463
	Barium	100	+/-100	U	100	P	04/16/2025	17:06	LB135463
	Cadmium	6.00	+/-6.00	U	6.00	P	04/16/2025	17:06	LB135463
	Chromium	10.0	+/-10.0	U	10.0	P	04/16/2025	17:06	LB135463
	Copper	20.0	+/-20.0	U	20.0	P	04/16/2025	17:06	LB135463
	Lead	12.0	+/-12.0	U	12.0	P	04/16/2025	17:06	LB135463
	Nickel	40.0	+/-40.0	U	40.0	P	04/16/2025	17:06	LB135463
	Selenium	20.0	+/-20.0	U	20.0	P	04/16/2025	17:06	LB135463
	Silver	10.0	+/-10.0	U	10.0	P	04/16/2025	17:06	LB135463

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT		SDG No.: Q1800							
Contract: ENTA05	Lab Code: CHEM	Case No.: Q1800	SAS No.: Q1800						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Zinc	40.0	+/-40.0	U	40.0	P	04/16/2025	17:06	LB135463
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	04/16/2025	17:59	LB135463
	Barium	100	+/-100	U	100	P	04/16/2025	17:59	LB135463
	Cadmium	6.00	+/-6.00	U	6.00	P	04/16/2025	17:59	LB135463
	Chromium	10.0	+/-10.0	U	10.0	P	04/16/2025	17:59	LB135463
	Copper	20.0	+/-20.0	U	20.0	P	04/16/2025	17:59	LB135463
	Lead	12.0	+/-12.0	U	12.0	P	04/16/2025	17:59	LB135463
	Nickel	40.0	+/-40.0	U	40.0	P	04/16/2025	17:59	LB135463
	Selenium	20.0	+/-20.0	U	20.0	P	04/16/2025	17:59	LB135463
	Silver	10.0	+/-10.0	U	10.0	P	04/16/2025	17:59	LB135463
	Zinc	40.0	+/-40.0	U	40.0	P	04/16/2025	17:59	LB135463
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	04/16/2025	18:52	LB135463
	Barium	100	+/-100	U	100	P	04/16/2025	18:52	LB135463
	Cadmium	6.00	+/-6.00	U	6.00	P	04/16/2025	18:52	LB135463
	Chromium	10.0	+/-10.0	U	10.0	P	04/16/2025	18:52	LB135463
	Copper	20.0	+/-20.0	U	20.0	P	04/16/2025	18:52	LB135463
	Lead	12.0	+/-12.0	U	12.0	P	04/16/2025	18:52	LB135463
	Nickel	40.0	+/-40.0	U	40.0	P	04/16/2025	18:52	LB135463
	Selenium	20.0	+/-20.0	U	20.0	P	04/16/2025	18:52	LB135463
	Silver	10.0	+/-10.0	U	10.0	P	04/16/2025	18:52	LB135463
	Zinc	40.0	+/-40.0	U	40.0	P	04/16/2025	18:52	LB135463
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	04/16/2025	19:43	LB135463
	Barium	100	+/-100	U	100	P	04/16/2025	19:43	LB135463
	Cadmium	6.00	+/-6.00	U	6.00	P	04/16/2025	19:43	LB135463
	Chromium	10.0	+/-10.0	U	10.0	P	04/16/2025	19:43	LB135463
	Copper	20.0	+/-20.0	U	20.0	P	04/16/2025	19:43	LB135463
	Lead	12.0	+/-12.0	U	12.0	P	04/16/2025	19:43	LB135463
	Nickel	40.0	+/-40.0	U	40.0	P	04/16/2025	19:43	LB135463
	Selenium	20.0	+/-20.0	U	20.0	P	04/16/2025	19:43	LB135463
	Silver	10.0	+/-10.0	U	10.0	P	04/16/2025	19:43	LB135463
	Zinc	40.0	+/-40.0	U	40.0	P	04/16/2025	19:43	LB135463
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	04/16/2025	20:13	LB135463
	Barium	100	+/-100	U	100	P	04/16/2025	20:13	LB135463
	Cadmium	6.00	+/-6.00	U	6.00	P	04/16/2025	20:13	LB135463
	Chromium	10.0	+/-10.0	U	10.0	P	04/16/2025	20:13	LB135463
	Copper	20.0	+/-20.0	U	20.0	P	04/16/2025	20:13	LB135463
	Lead	12.0	+/-12.0	U	12.0	P	04/16/2025	20:13	LB135463
	Nickel	40.0	+/-40.0	U	40.0	P	04/16/2025	20:13	LB135463
	Selenium	20.0	+/-20.0	U	20.0	P	04/16/2025	20:13	LB135463

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>ENTACT</u>				SDG No.:	<u>Q1800</u>			
Contract:	<u>ENTA05</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q1800</u>	SAS No.:	<u>Q1800</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Silver	10.0	+/-10.0	U	10.0	P	04/16/2025	20:13	LB135463
	Zinc	40.0	+/-40.0	U	40.0	P	04/16/2025	20:13	LB135463

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT **SDG No.:** Q1800
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1800 **SAS No.:** Q1800

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	------	---	------------------	------------------	---------------