

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: ENTACT **SDG No.:** Q1800
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1800 **SAS No.:** Q1800
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	-14.1			-20	20	04/16/2025	13:51	LB135463
	Barium	1.88	6.0	31	-94	106	04/16/2025	13:51	LB135463
	Cadmium	-1.99	1.0	199	-5	7	04/16/2025	13:51	LB135463
	Chromium	55.8	52.0	107	42	62	04/16/2025	13:51	LB135463
	Copper	1.52	2.0	76	-18	22	04/16/2025	13:51	LB135463
	Lead	-4.75			-12	12	04/16/2025	13:51	LB135463
	Nickel	1.69	2.0	84	-38	42	04/16/2025	13:51	LB135463
	Selenium	-3.36			-20	20	04/16/2025	13:51	LB135463
	Silver	-4.60			-10	10	04/16/2025	13:51	LB135463
	Zinc	4.25			-40	40	04/16/2025	13:51	LB135463
ICSAB01	Arsenic	94.0	104	90	88.4	120	04/16/2025	13:55	LB135463
	Barium	468	537	87	437	637	04/16/2025	13:55	LB135463
	Cadmium	1040	972	107	826	1120	04/16/2025	13:55	LB135463
	Chromium	554	542	102	460	624	04/16/2025	13:55	LB135463
	Copper	492	511	96	434	588	04/16/2025	13:55	LB135463
	Lead	43.2	49.0	88	37	61	04/16/2025	13:55	LB135463
	Nickel	995	954	104	810	1100	04/16/2025	13:55	LB135463
	Selenium	47.0	46.0	102	26	66	04/16/2025	13:55	LB135463
	Silver	179	201	89	170	232	04/16/2025	13:55	LB135463
	Zinc	1040	952	109	809	1095	04/16/2025	13:55	LB135463