

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q1800
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1800 **SAS No.:** Q1800

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167605BS							
Arsenic	ug/L	4000	4050		101	80 - 120	P
Barium	ug/L	1000	894		89	80 - 120	P
Cadmium	ug/L	1000	984		98	80 - 120	P
Chromium	ug/L	2000	1980		99	80 - 120	P
Copper	ug/L	1500	1520		101	80 - 120	P
Lead	ug/L	5000	4900		98	80 - 120	P
Nickel	ug/L	2500	2390		96	80 - 120	P
Selenium	ug/L	10000	10400		104	80 - 120	P
Silver	ug/L	380	367		97	80 - 120	P
Zinc	ug/L	1000	1010		101	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q1800
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1800 **SAS No.:** Q1800

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167607BS Mercury	ug/L	4.0	4.09		102	80 - 120	CV