

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1800 SAS No.: Q1800 SDG NO.: Q1800

Continuing Calib Date: 04/15/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 09:17 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.78	4.78	4.68	4.88	0.00
Aroclor-1242-2 (2)	4.80	4.80	4.70	4.90	0.00
Aroclor-1242-3 (3)	4.86	4.86	4.76	4.96	0.00
Aroclor-1242-4 (4)	4.98	4.98	4.88	5.08	0.00
Aroclor-1242-5 (5)	5.63	5.63	5.53	5.73	0.00
Tetrachloro-m-xylene	3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl	8.74	8.73	8.63	8.83	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1800 SAS No.: Q1800 SDG NO.: Q1800

Continuing Calib Date: 04/15/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 09:17 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1242-1	(1)	4.77	4.77	4.67	4.87	0.00
Aroclor-1242-2	(2)	4.79	4.78	4.68	4.88	-0.01
Aroclor-1242-3	(3)	4.96	4.96	4.86	5.06	0.00
Aroclor-1242-4	(4)	5.05	5.04	4.94	5.14	-0.01
Aroclor-1242-5	(5)	5.57	5.56	5.46	5.66	-0.01
Tetrachloro-m-xylene		3.69	3.68	3.58	3.78	-0.01
Decachlorobiphenyl		8.69	8.68	8.58	8.78	-0.01