

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1800 SAS No.: Q1800 SDG NO.: Q1800

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL01 Date Analyzed: 04/15/2025

Lab Sample No.: AR1660CCC500 Data File : PO110448.D Time Analyzed: 08:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.779	4.679	4.879	581.080	500.000	16.2
Aroclor-1016-2	4.798	4.698	4.898	584.320	500.000	16.9
Aroclor-1016-3	4.855	4.755	4.955	568.620	500.000	13.7
Aroclor-1016-4	4.976	4.875	5.075	582.890	500.000	16.6
Aroclor-1016-5	5.232	5.132	5.332	572.260	500.000	14.5
Aroclor-1260-1	6.273	6.172	6.372	582.680	500.000	16.5
Aroclor-1260-2	6.463	6.360	6.560	563.760	500.000	12.8
Aroclor-1260-3	6.831	6.729	6.929	558.960	500.000	11.8
Aroclor-1260-4	7.091	6.988	7.188	558.800	500.000	11.8
Aroclor-1260-5	7.333	7.230	7.430	556.980	500.000	11.4
Decachlorobiphenyl	8.737	8.632	8.832	45.490	50.000	-9.0
Tetrachloro-m-xylene	3.688	3.588	3.788	58.810	50.000	17.6

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1800 SAS No.: Q1800 SDG NO.: Q1800

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL01 Date Analyzed: 04/15/2025

Lab Sample No.: AR1660CCC500 Data File : PO110448.D Time Analyzed: 08:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.766	4.666	4.866	565.680	500.000	13.1
Aroclor-1016-2	4.785	4.684	4.884	564.260	500.000	12.9
Aroclor-1016-3	4.960	4.860	5.060	563.480	500.000	12.7
Aroclor-1016-4	5.003	4.902	5.102	516.050	500.000	3.2
Aroclor-1016-5	5.215	5.115	5.315	548.210	500.000	9.6
Aroclor-1260-1	6.247	6.146	6.346	549.300	500.000	9.9
Aroclor-1260-2	6.435	6.334	6.534	534.080	500.000	6.8
Aroclor-1260-3	6.588	6.487	6.687	531.770	500.000	6.4
Aroclor-1260-4	7.060	6.957	7.157	527.140	500.000	5.4
Aroclor-1260-5	7.300	7.197	7.397	531.100	500.000	6.2
Decachlorobiphenyl	8.688	8.584	8.784	48.740	50.000	-2.5
Tetrachloro-m-xylene	3.685	3.586	3.786	58.050	50.000	16.1