

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1800 SAS No.: Q1800 SDG NO.: Q1800

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL02 Date Analyzed: 04/15/2025

Lab Sample No.: AR1660CCC500 Data File : PO110463.D Time Analyzed: 16:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.781	4.679	4.879	574.340	500.000	14.9
Aroclor-1016-2	4.800	4.698	4.898	575.740	500.000	15.1
Aroclor-1016-3	4.856	4.755	4.955	560.830	500.000	12.2
Aroclor-1016-4	4.977	4.875	5.075	574.400	500.000	14.9
Aroclor-1016-5	5.234	5.132	5.332	552.080	500.000	10.4
Aroclor-1260-1	6.275	6.172	6.372	562.120	500.000	12.4
Aroclor-1260-2	6.465	6.360	6.560	550.100	500.000	10.0
Aroclor-1260-3	6.833	6.729	6.929	541.520	500.000	8.3
Aroclor-1260-4	7.093	6.988	7.188	522.350	500.000	4.5
Aroclor-1260-5	7.334	7.230	7.430	518.060	500.000	3.6
Decachlorobiphenyl	8.738	8.632	8.832	42.750	50.000	-14.5
Tetrachloro-m-xylene	3.690	3.588	3.788	57.130	50.000	14.3

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1800 SAS No.: Q1800 SDG NO.: Q1800

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL02 Date Analyzed: 04/15/2025

Lab Sample No.: AR1660CCC500 Data File : PO110463.D Time Analyzed: 16:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.767	4.666	4.866	564.770	500.000	13.0
Aroclor-1016-2	4.787	4.684	4.884	557.720	500.000	11.5
Aroclor-1016-3	4.962	4.860	5.060	561.810	500.000	12.4
Aroclor-1016-4	5.005	4.902	5.102	504.350	500.000	0.9
Aroclor-1016-5	5.217	5.115	5.315	538.570	500.000	7.7
Aroclor-1260-1	6.249	6.146	6.346	532.130	500.000	6.4
Aroclor-1260-2	6.437	6.334	6.534	521.520	500.000	4.3
Aroclor-1260-3	6.590	6.487	6.687	513.020	500.000	2.6
Aroclor-1260-4	7.062	6.957	7.157	500.790	500.000	0.2
Aroclor-1260-5	7.303	7.197	7.397	503.360	500.000	0.7
Decachlorobiphenyl	8.691	8.584	8.784	47.760	50.000	-4.5
Tetrachloro-m-xylene	3.687	3.586	3.786	57.040	50.000	14.1