

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1800 SAS No.: Q1800 SDG NO.: Q1800

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL01 Date Analyzed: 04/16/2025

Lab Sample No.: HSTDCCC750 Data File : PS029823.D Time Analyzed: 16:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	8.875	8.788	8.988	773.350	712.500	8.5
2,4-D	8.031	7.942	8.142	732.850	705.000	4.0
2,4-DCAA	6.953	6.863	7.063	779.070	750.000	3.9

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1800 SAS No.: Q1800 SDG NO.: Q1800

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL01 Date Analyzed: 04/16/2025

Lab Sample No.: HSTDCCC750 Data File : PS029823.D Time Analyzed: 16:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.553	9.451	9.651	756.220	712.500	6.1
2,4-D	8.672	8.571	8.771	686.520	705.000	-2.6
2,4-DCAA	7.476	7.375	7.575	736.660	750.000	-1.8