

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1800 SAS No.: Q1800 SDG NO.: Q1800

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL02 Date Analyzed: 04/16/2025

Lab Sample No.: HSTDCCC750 Data File : PS029833.D Time Analyzed: 20:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	8.874	8.788	8.988	758.730	712.500	6.5
2,4-D	8.030	7.942	8.142	720.210	705.000	2.2
2,4-DCAA	6.952	6.863	7.063	768.900	750.000	2.5

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1800 SAS No.: Q1800 SDG NO.: Q1800

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/02/2025 04/02/2025

Client Sample No.: CCAL02 Date Analyzed: 04/16/2025

Lab Sample No.: HSTDCCC750 Data File : PS029833.D Time Analyzed: 20:49

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.552	9.451	9.651	756.330	712.500	6.2
2,4-D	8.671	8.571	8.771	685.160	705.000	-2.8
2,4-DCAA	7.476	7.375	7.575	718.270	750.000	-4.2