

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041725\
 Data File : PS029837.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Apr 2025 12:55
 Operator : AR\AJ
 Sample : PB167608BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PB167608BS

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 04/18/2025
 Supervised By :mohammad ahmed 04/21/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 17 23:41:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040225.M
 Quant Title : 8080.M
 QLast Update : Wed Apr 02 23:52:55 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.951	7.473	1121.6E6	350.2E6	550.128	513.820
Target Compounds							
1) T	Dalapon	2.454	2.527	1842.8E6	723.0E6	543.210	455.406
2) T	3,5-DICHL...	6.153	6.467	1588.0E6	449.9E6	527.975	465.919
3) T	4-Nitroph...	6.741	7.007	733.8E6	339.1E6	522.385	478.225
5) T	DICAMBA	7.127	7.659	4466.4E6	1802.1E6	536.696	487.926
6) T	MCPD	7.304	7.765	264.4E6	75855996	48.779	44.926
7) T	MCPA	7.447	7.995	353.2E6	103.9E6	48.593m	45.660
8) T	DICHLORPROP	7.808	8.353	1115.1E6	464.3E6	506.985	480.604
9) T	2,4-D	8.029	8.669	1260.0E6	519.3E6	522.916	481.557
10) T	Pentachlo...	8.307	9.170	15826.5E6	9477.7E6	520.522	528.660
11) T	2,4,5-TP ...	8.872	9.549	6255.1E6	3757.7E6	538.521	520.959
12) T	2,4,5-T	9.155	9.955	6250.0E6	3538.0E6	543.265	527.841
13) T	2,4-DB	9.714	10.514	990.1E6	372.2E6	544.989	501.677
14) T	DINOSEB	10.878	10.888	4462.6E6	2605.5E6	528.621	509.594
15) T	Picloram	10.698	11.924	8102.7E6	5825.8E6	522.738	492.690m
16) T	DCPA	11.180	11.924	7712.3E6	4725.8E6	552.214	533.944m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS041725\
Data File : PS029837.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Apr 2025 12:55
Operator : AR\AJ
Sample : PB167608BS
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
PB167608BS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 04/18/2025
Supervised By : mohammad ahmed 04/21/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 17 23:41:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS040225.M
Quant Title : 8080.M
QLast Update : Wed Apr 02 23:52:55 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

