

RETENTION TIMES OF INITIAL CALIBRATION

Contract: ENTA05

Lab Code: CHEM Case No.: Q1800 SAS No.: Q1800 SDG NO.: Q1800

Instrument ID: ECD_L Calibration Date(s): 04/14/2025 04/14/2025

Calibration Times: 15:07 16:15

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID: RT 100 = PL095205.D RT 075 = PL095206.D

RT 050 = PL095207.D RT 025 = PL095208.D RT 005 = PL095209.D

COMPOUND	RT 100	RT 075	RT 050	RT 025	RT 005	MEAN RT	RT WINDOW	
							FROM	TO
Decachlorobiphenyl	9.05	9.05	9.05	9.05	9.05	9.05	8.95	9.15
Endrin	6.57	6.57	6.57	6.57	6.57	6.57	6.47	6.67
gamma-BHC (Lindane)	4.32	4.32	4.32	4.32	4.32	4.32	4.22	4.42
Heptachlor	4.91	4.91	4.91	4.91	4.91	4.91	4.81	5.01
Heptachlor epoxide	5.68	5.68	5.68	5.68	5.68	5.68	5.58	5.78
Methoxychlor	7.50	7.50	7.50	7.50	7.50	7.50	7.40	7.60
Tetrachloro-m-xylene	3.54	3.54	3.54	3.53	3.54	3.53	3.43	3.63



Contract:	<u>ENTA05</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1800</u>	SAS No.:	<u>Q1800</u>	SDG NO.:	<u>Q1800</u>
Instrument ID:	<u>ECD_L</u>	Calibration Date(s):		<u>04/14/2025</u>	<u>04/14/2025</u>		
		Calibration Times:		<u>15:07</u>	<u>16:15</u>		

LAB FILE ID:	RT 100 =	<u>PL095205.D</u>	RT 075 =	<u>PL095206.D</u>
RT 050 =	PL095207.D	RT 025 =	PL095208.D	RT 005 =
				PL095209.D

[illegible]