

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: ENTA05

Lab Code: CHEM **Case No.:** Q1800 **SAS No.:** Q1800 **SDG NO.:** Q1800

Instrument ID: ECD_L **Date(s) Analyzed:** 04/14/2025 04/14/2025

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	4.70	4.60	4.80	127558000
		2	5.23	5.13	5.33	130182000
		3	5.94	5.84	6.04	468201000
		4	6.02	5.92	6.12	550131000
		5	6.87	6.77	6.97	107924000
Toxaphene	500	1	6.23	6.13	6.33	27376000
		2	6.44	6.34	6.54	14916800
		3	7.06	6.96	7.16	82942000
		4	7.15	7.05	7.25	64931700
		5	7.93	7.83	8.03	46223600

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: ENTA05

Lab Code: CHEM **Case No.:** Q1800 **SAS No.:** Q1800 **SDG NO.:** Q1800

Instrument ID: ECD_L **Date(s) Analyzed:** 04/14/2025 04/14/2025

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Chlordane	500	1	3.76	3.66	3.86	147703000
		2	4.34	4.24	4.44	170474000
		3	4.97	4.87	5.07	498602000
		4	5.03	4.93	5.13	495465000
		5	5.93	5.83	6.03	183560000
Toxaphene	500	1	4.99	4.89	5.09	27499800
		2	5.32	5.22	5.42	25414900
		3	5.68	5.58	5.78	27465800
		4	6.59	6.49	6.69	96741800
		5	7.03	6.93	7.13	106909000