

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1800 SAS No.: Q1800 SDG NO.: Q1800

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL02 Date Analyzed: 04/16/2025

Lab Sample No.: PSTDCCC050 Data File : PL095266.D Time Analyzed: 16:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.054	8.953	9.153	57.190	50.000	14.4
Endrin	6.571	6.471	6.671	54.200	50.000	8.4
gamma-BHC (Lindane)	4.325	4.223	4.423	49.430	50.000	-1.1
Heptachlor	4.913	4.811	5.011	51.130	50.000	2.3
Heptachlor epoxide	5.681	5.580	5.780	52.540	50.000	5.1
Methoxychlor	7.499	7.398	7.598	49.060	50.000	-1.9
Tetrachloro-m-xylene	3.536	3.435	3.635	57.980	50.000	16.0

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1800 SAS No.: Q1800 SDG NO.: Q1800

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/14/2025 04/14/2025

Client Sample No.: CCAL02 Date Analyzed: 04/16/2025

Lab Sample No.: PSTDCCC050 Data File : PL095266.D Time Analyzed: 16:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.899	7.800	8.000	59.980	50.000	20.0
Endrin	5.627	5.527	5.727	54.280	50.000	8.6
gamma-BHC (Lindane)	3.599	3.499	3.699	48.420	50.000	-3.2
Heptachlor	3.936	3.836	4.036	50.980	50.000	2.0
Heptachlor epoxide	4.719	4.618	4.818	51.840	50.000	3.7
Methoxychlor	6.599	6.499	6.699	51.510	50.000	3.0
Tetrachloro-m-xylene	2.768	2.667	2.867	56.770	50.000	13.5