

## **Report of Analysis**

Client:	ENTACT			Date Collected:			
Project:	ooklyn, NY - E930	)9	Date Received:				
Client Sample ID:	PB167609BL			SDG No.:	Q1800		
Lab Sample ID:	PB167609BL			Matrix:	TCLP		
•				% Solid:	0	Decanted	
Analytical Method							:
Sample Wt/Vol:	1000 Units:	mL		Final Vol:	10000	uL	
Soil Aliquot Vol:		uL		Test:	TCLP Pestic	cide	
Extraction Type:				Injection Volum	e :		
GPC Factor :	1.0	PH :					
Prep Method :	3510C						
File ID/Qc Batch:	Dilution:	Prep D	ate	Date Analyzed	Pre	p Batch ID	
PL095257.D	1	04/16/2	25 08:45	04/16/25 13:46	PB	167609	
CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ/C	RQL	Units
TARGETS							
58-89-9	gamma-BHC (Lindane)	0.0000037	U	0.0000037		0.000050	mg/L
76-44-8	Heptachlor	0.0000027	U	0.0000027		0.000050	mg/L
1024-57-3	Heptachlor epoxide	0.0000096	U	0.0000096		0.000050	mg/L
72-20-8	Endrin	0.0000032	U	0.0000032		0.000050	mg/L
72-43-5	Methoxychlor	0.000011	U	0.000011		0.000050	mg/L
8001-35-2	Toxaphene	0.00017	U	0.00017		0.0010	mg/L
57-74-9	Chlordane	0.000088	U	0.000088		0.00050	mg/L
SURROGATES							
SURROGATES 2051-24-3 877-09-8	Decachlorobiphenyl Tetrachloro-m-xylene	20.6 18.9		30 (43) - 150 (140)		103%	SPK: 20 SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates > 25% difference for detected

concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

- J = Estimated Value
- B = Analyte Found in Associated Method Blank
- N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration

was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit