

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1800

Client: ENTACT

Analytical Method: 8081B

DataFile : PL095263.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Client Sample ID: Q1800-03MS	WC-A4-01-CMS											
	gamma-BHC (Lindane)	5	0	5.30	ug/L	106				30 (60)	150 (152)	
	Heptachlor	5	0	5.10	ug/L	102				30 (56)	150 (147)	
	Heptachlor epoxide	5	0	5.40	ug/L	108				30 (77)	150 (143)	
	Endrin	5	0	5.90	ug/L	118				30 (76)	150 (144)	
	Methoxychlor	5	0	5.50	ug/L	110				30 (70)	150 (142)	