

Surrogate Summary

SW-846

SDG No.: Q1800

Client: ENTACT

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167587TB	PB167587TB	2-Fluorophenol	150	129	86		15 (10)	110 (139)
		Phenol-d6	150	124	83		15 (10)	110 (134)
		Nitrobenzene-d5	100	91.8	92		30 (49)	130 (133)
		2-Fluorobiphenyl	100	91.0	91		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	142	95		15 (44)	110 (137)
		Terphenyl-d14	100	112	112		30 (48)	130 (125)
PB167606BL	PB167606BL	2-Fluorophenol	150	133	89		15 (10)	110 (139)
		Phenol-d6	150	119	80		15 (10)	110 (134)
		Nitrobenzene-d5	100	89.0	89		30 (49)	130 (133)
		2-Fluorobiphenyl	100	90.1	90		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	118	79		15 (44)	110 (137)
		Terphenyl-d14	100	86.7	87		30 (48)	130 (125)
PB167606BS	PB167606BS	2-Fluorophenol	150	135	90		15 (10)	110 (139)
		Phenol-d6	150	123	82		15 (10)	110 (134)
		Nitrobenzene-d5	100	85.2	85		30 (49)	130 (133)
		2-Fluorobiphenyl	100	85.9	86		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	124	83		15 (44)	110 (137)
		Terphenyl-d14	100	90.2	90		30 (48)	130 (125)
Q1787-12MS	TP-1MS	2-Fluorophenol	150	124	83		15 (10)	110 (139)
		Phenol-d6	150	103	68		15 (10)	110 (134)
		Nitrobenzene-d5	100	95.4	95		30 (49)	130 (133)
		2-Fluorobiphenyl	100	88.7	89		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	125	83		15 (44)	110 (137)
		Terphenyl-d14	100	89.1	89		30 (48)	130 (125)
Q1787-12MSD	TP-1MSD	2-Fluorophenol	150	130	87		15 (10)	110 (139)
		Phenol-d6	150	109	72		15 (10)	110 (134)
		Nitrobenzene-d5	100	96.9	97		30 (49)	130 (133)
		2-Fluorobiphenyl	100	90.8	91		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	133	88		15 (44)	110 (137)
		Terphenyl-d14	100	96.6	97		30 (48)	130 (125)
Q1800-03	WC-A4-01-C	2-Fluorophenol	150	115	76		15 (10)	110 (139)
		Phenol-d6	150	104	69		15 (10)	110 (134)
		Nitrobenzene-d5	100	88.0	88		30 (49)	130 (133)
		2-Fluorobiphenyl	100	83.0	83		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	136	91		15 (44)	110 (137)
		Terphenyl-d14	100	94.0	94		30 (48)	130 (125)