

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1800 SAS No.: Q1800 SDG NO.: Q1800

EPA Sample No.: SSTDCCC040 Date Analyzed: 04/16/2025

Lab File ID: BM049944.D Time Analyzed: 09:28

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	390253	7.769	1356600	10.56	881748	14.42
UPPER LIMIT	780506	8.269	2713200	11.063	1763500	14.916
LOWER LIMIT	195127	7.269	678300	10.063	440874	13.916
EPA SAMPLE NO.						
01 PB167587TB	340959	7.77	1139920	10.56	732723	14.42

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.