

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1800 SAS No.: Q1800 SDG NO.: Q1800

EPA Sample No.: SSTDCCC040 Date Analyzed: 04/16/2025

Lab File ID: BP024304.D Time Analyzed: 09:43

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	273289	7.728	1112650	10.50	684984	14.36
UPPER LIMIT	546578	8.228	2225300	10.999	1369970	14.857
LOWER LIMIT	136645	7.228	556325	9.999	342492	13.857
EPA SAMPLE NO.						
01 PB167606BL	275099	7.73	1054720	10.51	594014	14.36
02 PB167606BS	272135	7.73	1102100	10.50	648559	14.36
03 WC-A4-01-C	292453	7.73	1142810	10.50	666880	14.36

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.