

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1800

Client: ENTACT

Analytical Method: SW8260-Low

Datafile : VX045801.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0416WBSD01	Vinyl chloride	20	17.2	ug/L	86	3		70 (65)	130 (117)	20 (20)
	1,1-Dichloroethene	20	18.0	ug/L	90	2		70 (74)	130 (110)	20 (20)
	2-Butanone	100	110	ug/L	110	11		40 (65)	160 (122)	20 (20)
	Carbon Tetrachloride	20	19.8	ug/L	99	1		70 (77)	130 (113)	20 (20)
	Chloroform	20	20.0	ug/L	100	1		70 (79)	130 (113)	20 (20)
	Benzene	20	19.1	ug/L	96	1		70 (82)	130 (109)	20 (20)
	1,2-Dichloroethane	20	21.0	ug/L	105	3		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	18.6	ug/L	93	4		70 (77)	130 (113)	20 (20)
	Tetrachloroethene	20	19.4	ug/L	97	1		70 (67)	130 (123)	20 (20)
	Chlorobenzene	20	19.7	ug/L	99	2		70 (82)	130 (109)	20 (20)