

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q1800 SAS No.: Q1800 SDG NO.: Q1800
 Lab File ID: VX045797.D Date Analyzed: 04/16/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:28
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	85794	5.54	152447	6.75	137129	10.05
UPPER LIMIT	171588	6.038	304894	7.251	274258	10.549
LOWER LIMIT	42897	5.038	76223.5	6.251	68564.5	9.549
EPA SAMPLE NO.						
WC-A4-01-G	66619	5.54	132587	6.76	124254	10.05
VX0416WBL01	72471	5.54	141858	6.76	129459	10.05
VX0416WBS01	91256	5.54	163123	6.76	144629	10.05
VX0416WBSD01	87856	5.54	155557	6.76	137506	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.