

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GFEL01

Lab Code: CHEM Case No.: Q1802 SAS No.: Q1802 SDG No.: Q1802

Instrument ID: MSVOA_L Calibration Date/Time: 04/15/2025 11:30

Lab File ID: VL042326.D Init. Calib. Date(s): 03/27/2025 03/27/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 08:26 11:36

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Vinyl Chloride	0.639	0.483		-24.41	30
1,1-Dichloroethane	1.183	0.983		-16.91	30
cis-1,2-Dichloroethene	1.423	1.208		-15.11	30
1,1,1-Trichloroethane	2.349	1.810		-22.95	30
Trichloroethene	0.414	0.377		-8.94	30
Tetrachloroethene	0.393	0.316		-19.59	30
1-Bromo-4-Fluorobenzene	0.821	0.792		-3.53	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.