

### Initial and Continuing Calibration Blank Summary

**Client:** Scalamandre – Tully JV

**SDG No.:** Q1803

**Project:** NYC DOT Harper Street Yard North

**RunNo.:** LB135440

| Analyte                                    | Units | Result | Acceptance Limits | Conc Qual | MDL     | RDL   | Analysis Date |
|--------------------------------------------|-------|--------|-------------------|-----------|---------|-------|---------------|
| Sample ID: <b>ICB1</b><br>Reactive Cyanide | mg/L  | 0.0014 | 0.0025            | J         | 0.00096 | 0.005 | 04/15/2025    |
| Sample ID: <b>CCB1</b><br>Reactive Cyanide | mg/L  | 0.0014 | 0.0025            | J         | 0.00096 | 0.005 | 04/15/2025    |
| Sample ID: <b>CCB2</b><br>Reactive Cyanide | mg/L  | 0.0012 | 0.0025            | J         | 0.00096 | 0.005 | 04/15/2025    |
| Sample ID: <b>CCB3</b><br>Reactive Cyanide | mg/L  | 0.0013 | 0.0025            | J         | 0.00096 | 0.005 | 04/15/2025    |
| Sample ID: <b>CCB4</b><br>Reactive Cyanide | mg/L  | 0.0015 | 0.0025            | J         | 0.00096 | 0.005 | 04/15/2025    |